

INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH

An Autonomous Institution, Ministry of Education, Govt. of India

Dr. Homi Bhabha Road, Pashan Pune – 411008.

Tel: +91-020-25898017; Email: purchase@iiser.pune.ac.in

Website: www.iiserpune.ac.in

DATE: 04/04/2022

DOMESTIC TENDER REFERENCE NO: DOMESTIC/IISER/PUR/1431/21


Tender ID: 2022 IISRP 676584 1

CORRIGENDUM

With tender reference no DOMESTIC/IISER/PUR/1431/21 dated 03/03/2022, for Establishing a Biosafety Level 3 (BSL 3) Laboratory at IISER Pune. It is clarified that tender quantity / specifications amended as below:

Tendered / Pre-Bid Nomenclature	Amendment in Nomenclature
<p>The refrigerators/freezers must have the following capacities: -80C freezer (qty: 1) > 375L; -40C freezer (qty: 1) > 540L; -40C freezer (qty: 1) > 250L; 4C refrigerator (qty: 1) > 640L; 4C refrigerator (qty: 1) > 375L</p> <p>The CO2 incubators must have a capacity of > 180L. Qty needed for this is 2 (two).</p> <p>The centrifuge must have a capacity of > 1.5L, and must come with all necessary accessories (e.g. buckets, adaptors, rotors) capable of centrifuging 50 mL, 15 mL and 1.5 mL tubes.</p> <p>Since this is a turnkey project, other specifications may be determined by the bidder, such that they comply with the design of the BSL3 facility as per guidelines provided by DBT.</p> <p>The freezers, refrigerators, CO2 incubators and centrifuge must be supplied from the approved make list only. No other vendors or</p>	<p><i>The refrigerators/freezers must have the following capacities: -80C freezer (qty: 1) > 375L; -20C freezer (qty: 1) > 250L; 4C refrigerator (qty: 1) > 375L</i></p> <p><i>The CO2 incubators must have a capacity of > 180L. Qty needed for this is 1 (one).</i></p> <p><i>The centrifuge (qty: 1) must have a capacity of > 1.5L, and must come with all necessary accessories (e.g. buckets, adaptors, rotors) capable of centrifuging 50 mL, 15 mL and 1.5 mL tubes.</i></p> <p><i>Since this is a turnkey project, other specifications may be determined by the bidder, such that they comply with the design of the BSL3 facility as per guidelines provided by DBT.</i></p> <p><i>The freezers, refrigerators, CO2 incubators and centrifuge must be supplied from the</i></p>

make will be allowed.	<i>approved make list only. No other vendors or make will be allowed.</i>
Only Class II B2 biosafety cabinets should be installed. We require at least 6 Class II B2 biosafety cabinets. No class II A2 are required.	<i>Only Class II B2 biosafety cabinets should be installed. We require 2 Class II B2 biosafety cabinets.</i> <i>No class II A2 are required.</i>



Assistant Registrar (S&P)