



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH PUNE

PREBID CLARIFICATION ON TENDER NUMBER - IISER/PUR/1749/21

Item Description- Procurement of Fully Motorized High-End Operational Fluorescence Live Cell Workstation.

Refer an open tender published on Institute website www.iiserpune.ac.in and on CPP Portal on **25/04/2022**.

Pre-Bid meeting was held on **02/05/2022** at **3.00 PM** and minutes of meeting is as under:

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the tender and thereafter requested Assistant Registrar (S&P) to brief the bidders on the salient features of the tender.

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website www.iiserpune.ac.in will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

02/05/2022

Sd/-
Assistant Registrar (S&P)

IISER PUNE
PRE-BID CONFERENCE FOR PROCUREMENT OF FULLY MOTORIZED HIGH-END OPERATIONAL FLUORESCENCE
LIVE CELL WORKSTATION

IISER/PUR/1749/21
TECHNICAL QUERIES AND CLARIFICATION

S. No	Query/Clarification Sought	Clarification / Amendment
1	Stage warmer: Do we need to offer the CO2 incubator or only stage warmer is sufficient?	CO2 incubator is not required.
2	Fast emission filter wheel is asked, kindly do let us know if you have any preference for the emission filter positions	Emission filter wheel with 6-8 positions should to be quoted.
3	Image splitter capability is asked as a software feature. Do you need us to offer the physical splitter or just the software be capable of image splitting and streaming two colour image as two separate channels?	Yes, a physical splitter should be offered. Please refer to the technical specifications in the tender. The same is already included.

IISER PUNE
PRE-BID CONFERENCE FOR PROCUREMENT OF FULLY MOTORIZED HIGH-END OPERATIONAL FLUORESCENCE
LIVE CELL WORKSTATION

IISER/PUR/1749/21
COMMERCIAL QUERIES AND CLARIFICATION

S. No	Query/Clarification Sought	Clarification / Amendment
1.	Please provide us option in the BOQ to quote optional items	Bidder may attach pdf document file for optional items with price details. The bidder should quote price of main equipment in the excel BOQ. However L1 will be determined by comparing the main equipment cost as per excel BOQ.
2.	EMD Amount - Rs 1,60,000/- Please clarify the EMD amount, it is mentioned as Rs 90,000/- on page no 45 under check list for Bidders	EMD Amount - Rs 1,60,000/- Please read EMD amount as Rs 1,60,000/- on page no 45 under check list for Bidders