



## INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH PUNE

### **PREBID CLARIFICATION ON TENDER NUMBER - DOMESTIC/IISER/PUR/0010/22**

#### **Item Description- Procurement of Confocal Micro Raman Spectrometer**

Refer an open tender published on Institute website [www.iiserpune.ac.in](http://www.iiserpune.ac.in) and on CPP Portal on **23/05/2022**.

Pre-Bid meeting was held on **30/05/2022** at **3.00 PM** and minutes of meeting is as under:

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the tender and thereafter requested Assistant Registrar (S&P) to brief the bidders on the salient features of the tender.

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website [www.iiserpune.ac.in](http://www.iiserpune.ac.in) will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

**30/05/2022**

**Sd/-  
Assistant Registrar (S&P)**

**IISER PUNE**  
**PRE-BID CONFERENCE FOR PROCUREMENT OF CONFOCAL MICRO RAMAN SPECTROMETER**

**DOMESTIC/IISER/PUR/0010/22**

**TECHNICAL QUERIES AND CLARIFICATION**

| S. No | Query/Clarification Sought  | Clarification / Amendment  |
|-------|---|--|
| 1     | <p><b><u>Chapter No.4 Page No. 21 Unit-Raman Spectrometer Specification Amended as:</u></b></p> <p>Spectral calibration should be automated. Vendor should provide the specific details in the technical documents if the calibration reference is built-in with the system</p>                       | <p><b><u>Chapter No.4 Page No. 21 Unit-Raman Spectrometer Specification Amended as:</u></b></p> <p>Vendor should supply reference calibration sample (such as silicon) and calibration lamp (like Neon). The lamp can also stay as built-in to the system. The calibration should be automated.</p>          |
| 2     | <p><b><u>Chapter No. 4 Page No. 21 Unit-CCD Detector Specification Amended as:</u></b></p> <p>Minimal 1650 x 200 or 1024x256 pixel format with Peltier cooling down to -55 deg C or better,<br/> Peak QE &gt; 50 %,<br/> It must have fully automated operation for providing maximal performance</p> | <p><b><u>Chapter No. 4 Page No. 21 Unit-CCD Detector Specification Amended as:</u></b></p> <p>Minimal 1650 x 200 or 1024x256 pixel format with Peltier cooling down to -55 deg C or better,<br/> <b>Peak QE &gt; 40 %</b>,<br/> It must have fully automated operation for providing maximal performance</p> |