



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH PUNE

PREBID CLARIFICATION ON TENDER NUMBER - DOMESTIC/IISER/PUR/0694/22

Item Description- Procurement of Near-Field Optical Microscopy.

Refer a domestic tender published on Institute website www.iiserpune.ac.in and on CPP Portal on **21/09/2022**.

Pre-Bid meeting was held on **29/09/2022** at **3.00 PM** and minutes of meeting is as under:

At the outset, the committee welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the tender and thereafter requested Assistant Registrar (S&P) to brief the bidders on the salient features of the tender.

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website www.iiserpune.ac.in will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

29/09/2022

Sd/-
Assistant Registrar (S&P)

TECHNICAL & COMMERCIAL QUERIES AND CLARIFICATION**PRE-BID CONFERENCE FOR PROCUREMENT OF NEAR-FIELD OPTICAL MICROSCOPY**

S.No	Query/Clarification Sought	Clarification / Amendment
1	<u>Page No. 20 Chapter No.4 Sr. No.1:</u> The desired resolution (below 1 nm) might be difficult to achieve.	<u>Page No. 20 Chapter No.4 Sr. No.1 amended as:</u> Resolution 5 nm or better is preferred.
2	What is the power requirements of the three lasers	Power of the lasers should be such that the instrument meet the resolution requirements and SNR.