



## भारतीय विज्ञान शिक्षा एवं अनुसंधान संस्थान पुणे

INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH PUNE

निविदा संख्या पर स्पष्टीकरण - आईआईएसईआर/पीयूआर/0522/23

CLARIFICATION ON TENDER NUMBER - IISER/PUR/0522/23

### वस्तु विवरण- कंटिन्युस फ्लो क्रयोस्टाट

#### ITEM DESCRIPTION- CONTINUOUS FLOW CRYOSTAT

कंटिन्युस फ्लो क्रयोस्टाट सिस्टम फॉर न्यारो गैप मनेट अप्लिकेशन की खरीद के लिए IISER पुणे निविदा संख्या IISER/PUR/0522/23 दिनांक 04/08/2023 देखें।

Refer IISER Pune tender number no. IISER/PUR/0522/23 dated 04/08/2023 for Procurement of Continuous Flow Cryostat.

प्री-बिड मीटिंग 11/08/2023 को अपराह्न 3.00 बजे आयोजित की गई थी।

Pre-Bid meeting was held on 11/08/2023 at 3.00 PM.

प्रारंभ में, समिति ने सभी सदस्यों और संभावित बोलीदाताओं के प्रतिनिधियों का स्वागत किया और सामान्य तौर पर निविदा के दायरे की जानकारी दी और उसके बाद कार्यालय अधीक्षक (एसएंडपी) से बोलीदाताओं को निविदा की मुख्य विशेषताओं के बारे में जानकारी देने का अनुरोध किया।

At the outset, the committee welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the tender and thereafter requested Office Superintendent (S&P) to brief the bidders on the salient features of the tender.

उपस्थित प्रतिनिधि दिए गए उत्तरों से संतुष्ट थे और यह सूचित किया गया था कि प्री-बिड कॉन्फ्रेंस के दौरान की गई चर्चा के अनुसार दिए गए सुधार / परिवर्धन / स्पष्टीकरण को IISER पुणे की वेबसाइट पर होस्ट किया जाएगा और सभी संभावित बोलीदाताओं को बोली दस्तावेजों में निर्धारित अनुसार अपनी बोली जमा करने से पहले प्री-बिड सम्मेलन की कार्यवाही का संज्ञान लेना आवश्यक है।

The representatives present were satisfied with the replies given and it was informed that the corrections / additons / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

हमारी आईआईएसईआर वेबसाइट [www.iiserpune.ac.in](http://www.iiserpune.ac.in) पर जारी नोटिस के अन्य नियम और शर्तें अपरिवर्तित रहेंगी। इस संबंध में और कोई पत्राचार नहीं किया जाएगा

The other terms & conditions of the notice issued on our IISER website [www.iiserpune.ac.in](http://www.iiserpune.ac.in) will remain unchanged. No more correspondence in this regard will be entertained

बैठक अध्यक्ष के धन्यवाद प्रस्ताव के साथ समाप्त हुई।

The meeting ended with vote of thanks to the Chair

11/08/2023

11/08/2023

सहायक कुलसचिव (भांडार एवं क्रय)

Assistant Registrar (S&P)

**TECHNICAL AND COMMERCIAL QUERIES AND CLARIFICATION****PRE-BID CONFERENCE FOR PROCUREMENT OF CONTINUOUS FLOW CRYOSTAT**

S. No	Query/Clarification Sought	Clarification / Amendment
1	As per BQ single price to be offered. Please confirm if we have to submit the price with main system price or we have to add PDF file for Accessories options.	Quote as a single price but show price break-up in pdf format of the detailed breakup of price bid packet.
2	As per tender doc. Delivery to be offered 60 days. Due to shortage of components and late delivery schedule, we request you to kindly accept delivery period 8 months after receipt of LC.	Delivery period is amended from 60 days to 8 Months.