

ADDENDUM TO PRE-BID CLARIFICATION ON TENDER NUMBER - IISER/PUR/0780/23

ANNEXURE -II

PROCUREMENT OF SPINNING DISK CONFOCAL MICROSCOPE

S. No	Query/Clarification Sought	Clarification / Amendment
1	2D and 3D blind deconvolution module should be quoted with super-resolution, deconvolution module for SORA spinning disk images. Super-resolution deconvolution should be capable of achieving the 120 nm or better resolution in XY and 320nm or better resolution in Z. The super resolution should be achieved by psf fitting algorithm for spinning disk and should be able to perform during acquisition and post-acquisition.	<u>Tender specification is amended to</u> 2D and 3D deconvolution modules should be quoted along with the super-resolution module. Super-resolution module should be capable of achieving 120nm or better resolution in xy and 350nm or better resolution in z.
2	We have gone through the BOQ, there is no provision to quote optional items. There are few items in the tender that needs to be quoted as optional. We request you to kindly make this provision in BOQ.	There is already a provision in place. In addition to the main equipment, the bidder may also submit price bids for optional items in the finance cover's designated pdf format (detailed breakup of price bid with bill of material).

Sd/-

सहायक कुलसचिव (भांडार एवं क्रय)
Assistant Registrar (S & P)