



भारतीय विज्ञान शिक्षा एवं अनुसंधान संस्थान पुणे
INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH PUNE
IISER/PUR/1625/23
PRE-BID CLARIFICATION

Item Description - Procurement of Field Emission Scanning Electron Microscope (FESEM) with integrated EDS, EBSD, and EBL attachments

एकीकृत ईडीएस, ईबीएसडी और ईबीएल अटैचमेंट के साथ फील्ड एमिशन स्कैनिंग इलेक्ट्रॉन माइक्रोस्कोप (एफईएसईएम) की खरीद के लिए संस्थान की वेबसाइट www.iiserpune.ac.in और सीपीपी पोर्टल पर 14/06/2024 को प्रकाशित वैश्विक निविदा देखें।

Refer to a Global tender published on the Institute website www.iiserpune.ac.in and on the CPP Portal on 14/06/2024 for procurement of Field Emission Scanning Electron Microscope (FESEM) with integrated EDS, EBSD, and EBL attachments.

बोली-पूर्व बैठक 26/06/2024 को अपराह्न 3.00 बजे आयोजित की गई और बैठक का विवरण इस प्रकार है:

Pre-Bid meeting was held on 26/06/2024 at 3.00 PM and minutes of meeting is as under:

प्रारंभ में, समिति ने सभी सदस्यों और संभावित बोलीदाताओं के प्रतिनिधियों का स्वागत किया और सामान्य तौर पर निविदा के दायरे की जानकारी दी और उसके बाद सहायक कुलसचिव (भांडार एवं क्रय) से बोली लगाने वालों को निविदा की मुख्य विशेषताओं के बारे में जानकारी देने का अनुरोध किया।

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the tender and thereafter requested Assistant Registrar (S&P) to brief the bidders on the salient features of the tender.

उपस्थित प्रतिनिधि दिए गए उत्तरों से संतुष्ट थे और यह सूचित किया गया था कि प्री बिड कॉन्फ्रेंस के दौरान की गई चर्चा के अनुसार दिए गए सुधार- / परिवर्धन / स्पष्टीकरण क IISER पुणे की वेबसाइट पर होस्ट किया जाएगा और सभी संभावित बोलीदाताओं को बोली दस्तावेजों में निर्धारित अनुसार अपनी बोली जमा करने से पहले प्री-बिड सम्मेलन की कार्यवाही का संज्ञान लेना आवश्यक है।

The representatives present were satisfied with the replies given and it was informed that the corrections / additons / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

हमारी आईआईएसईआर वेबसाइट www.iiserpune.ac.in और CPP portal पर जारी नोटिस के अन्य नियम और शर्तें अपरिवर्तित रहेंगी। इस संबंध में और कोई पत्राचार नहीं किया जाएगा।

The other terms & conditions of the notice issued on our IISER website www.iiserpune.ac.in and CPP portal will remain unchanged. No more correspondence in this regard will be entertained.

बैठक अध्यक्ष के धन्यवाद प्रस्ताव के साथ समाप्त हुई।

The meeting ended with vote of thanks to the Chair.

01/07/2024

Sd/-
सहायक कुलसचिव (भांडारण एवं क्रय)
Assistant Registrar (S&P)

TECHNICAL AND COMMERCIAL QUERIES AND CLARIFICATION**PRE-BID CONFERENCE FOR PROCUREMENT OF FIELD EMISSION SCANNING ELECTRON MICROSCOPE (FESEM) WITH INTEGRATED EDS, EBSD, AND EBL ATTACHMENTS**

S. No.	Query/Clarification Sought	Clarification / Amendment
1	<p>The offer should include the provision for replacing the faulty filament up to 10 years (including all the accessories required) after the warranty period subject to a maximum of two replacements.</p> <p><u>Requested</u> the price should be shown separately</p>	Tender specification prevails
2	<p>In Lens SE detector: column mounted, scintillator detector with optically coupled photomultiplier.</p> <p><u>Requested:</u> Scintillator detector with optically coupled photomultiplier</p>	<p><u>Original:</u> InLens SE detector: column mounted, scintillator detector with optically coupled photomultiplier</p> <p><u>Amended:</u> InLens SE detector</p>
3	<p>Integrated Air lock Mechanism will be preferred</p> <p><u>Requested</u> the price should be shown separately</p>	Tender specification prevails
4	<p>Vendors should have supplied at least five similar FESEMs in India in the last 5 years. Include the list of last five years FESEM installations in India with names, addresses and contact information of the end users in various Indian universities / institutes / laboratories should be provided.</p> <p><u>Requested:</u> We request you to change “Vendors should have supplied at least Four similar FESEMs in India in previously.</p>	Tender specification prevails
5	<p>Resolution specification as per the tender SI. No: 2 page 21 of 53: 0.9nm or better at 15kV, request to change as 0.7nm or better at 15kV</p>	Tender specification prevails

6	Electron optics as per the tender Sl. No: 6 (page 22 of 53): Hybrid magnetic/electrostatic objective/compound lens, request to change as Hybrid magnetic & electrostatic objective/compound lens.	Tender specification prevails
7	As per the tender document warranty is 3 years as per page 16 of 53 and warranty of 2 years as per page 25 of 53.	Clarification: the information as per page 25 of 53 is correct. In the amended document, this has been corrected on page 16 of 53.
8	<p><u>Page No.18, Clause No.16.2</u></p> <p>Penalty for delayed Services / LD:</p> <p>If the supplier fails to Supply, Install and Commission the system as per specifications mentioned in the order within the due date, the Supplier is liable to pay liquidated damages of 0.5 of order value per every week of delay subject to a maximum of 5% beyond the due date. Such money will be deducted from any amount due or which may become due to the supplier.</p>	<p><u>Page No.18, Clause No.16.2</u></p> <p>Clause in respect of Penalty for delayed Services / LD: is modified as below.</p> <p>If the supplier fails to Supply, Install and Commission the system as per specifications mentioned in the order within the due date, the Supplier is liable to pay liquidated damages of 0.5% of order value per every week of delay subject to a maximum of 10% beyond the due date. Such money will be deducted from any amount due or which may become due to the supplier.</p>