



भारतीय विज्ञान शिक्षा एवं अनुसंधान संस्थान पुणे
INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH PUNE
IISER/PUR/0894/24
PREBID CLARIFICATION ON GeM TENDER NUMBER - GEM/2023/B/5419892

कार्डियोवास्कुलर लैब के लिए हेमोडायनामिक रिकॉर्डर की खरीद
Item Description- Procurement of Hemodynamic Recorder for Cardiovascular Lab

कार्डियोवास्कुलर लैब के लिए हेमोडायनामिक रिकॉर्डर की खरीद के लिए संस्थान की वेबसाइट www.iiserpune.ac.in और GeM पोर्टल पर 26/09/2024 को प्रकाशित एक खुली निविदा देखें।

Refer to an open tender published on the Institute website www.iiserpune.ac.in and on the GeM Portal on 26/09/2024 for procurement of Hemodynamic Recorder for Cardiovascular Lab.

बोली-पूर्व बैठक 03/10/2024 को अपराह्न 3.00 बजे आयोजित की गई और बैठक का विवरण इस प्रकार है:
Pre-Bid meeting was held on 03/10/2024 at 3.00 PM and minutes of meeting is as under:

प्रारंभ में, समिति ने सभी सदस्यों और संभावित बोलीदाताओं के प्रतिनिधियों का स्वागत किया और सामान्य रूप से निविदा के दायरे के बारे में जानकारी दी और उसके बाद सहायक रजिस्ट्रार (एसएंडपी) से बोलीदाताओं को निविदा की मुख्य विशेषताओं के बारे में जानकारी देने का अनुरोध किया।

At the outset, the committee welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the tender and thereafter requested Assistant Registrar (S&P) to brief the bidders on the salient features of the tender.

उपस्थित प्रतिनिधि दिए गए उत्तरों से संतुष्ट थे और यह सूचित किया गया कि दिए गए सुधार / परिवर्धन / स्पष्टीकरण, जैसा कि प्री-बिड कॉन्फ्रेंस के दौरान चर्चा की गई थी, आईआईएसईआर पुणे की वेबसाइट पर होस्ट किए जाएंगे और सभी संभावित बोलीदाताओं को इसका संज्ञान लेना होगा। बोली दस्तावेजों में निर्धारित अनुसार अपनी बोलियां जमा करने से पहले बोली-पूर्व सम्मेलन की कार्यवाही।

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

हमारी आईआईएसईआर वेबसाइट www.iiserpune.ac.in और GeM portal पर जारी नोटिस के अन्य नियम और शर्तें अपरिवर्तित रहेंगी। इस संबंध में और कोई पत्राचार नहीं किया जाएगा।

The other terms & conditions of the notice issued on our IISER website www.iiserpune.ac.in and GeM portal will remain unchanged. No more correspondence in this regard will be entertained.

बैठक अध्यक्ष के धन्यवाद प्रस्ताव के साथ समाप्त हुई।
The meeting ended with vote of thanks to the Chair.

09/10/2024
09/10/2024

Sd/-
सहायक कुलसचिव (भंडारण एवं क्रय)
Assistant Registrar (S&P)

TECHNICAL AND COMMERCIAL QUERIES AND CLARIFICATION

PRE-BID CONFERENCE FOR ITEM DESCRIPTION- PROCUREMENT OF HEMODYNAMIC RECORDER FOR CARDIOVASCULAR LAB

S. No.	Query/Clarification Sought	Clarification / Amendment
1	<p><u>GeM Bid Document, Page No: 4, Buyer specification document</u></p> <p>The phrase "<i>Catheter Interface Cable for connection to any ADI Bridge Amp</i>" should be modified to "<i>Catheter Interface Cable for connection to any Bridge Amp</i>" by removing the word "ADI." This will make the specification more generalized.</p>	<p><u>GeM Bid Document, Page No: 4, Buyer specification document</u></p> <p>The specification is amended from Catheter Interface Cable for connection to any ADI Bridge Amp</p> <p>To</p> <p><i>Catheter Interface Cable for connection to any Bridge Amp</i></p>
2	<p><u>GeM Bid Document, Page No: 4, Buyer specification document</u></p> <p>Software Requirement: The sentence "<i>Software should support real-time streaming of raw data to MatLab and Python</i>" should be updated to "<i>Software should allow OLE linking of streaming raw data to MatLab</i>" by removing the word "Python."</p>	<p><u>GeM Bid Document, Page No: 4, Buyer specification document</u></p> <p>The specification is amended from Software should support real time streaming of raw data to MatLab and Python</p> <p>To</p> <p>Software should allow OLE linking of streaming raw data to MATLAB</p>