



**भारतीय विज्ञान शिक्षा एवं अनुसंधान संस्थान पुणे**  
**INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH PUNE**  
**IISER/PUR/2085/24**  
**PREBID CLARIFICATION**

**ITEM DESCRIPTION- PROCUREMENT OF FOUR PORTS GLOVE BOX**

यूवी विज़िबल डिटेक्टर के साथ एचपीएलसी प्रणाली की खरीद के लिए संस्थान की वेबसाइट [www.iiserpune.ac.in](http://www.iiserpune.ac.in) और सीपीपी पोर्टल पर 24/02/2025 को प्रकाशित खुली निविदा देखें।

Refer to an open tender published on the Institute website [www.iiserpune.ac.in](http://www.iiserpune.ac.in) and on the GeM Portal on 24/02/2025 for procurement of Four Ports Glove Box.

प्री-बिड मीटिंग 05/03/2025 को शाम 4.00 बजे आयोजित की गई और बैठक का कार्यवृत्त निम्नानुसार है:  
Pre-Bid meeting was held on 05/03/2025 at 4.00 PM and minutes of meeting is as under:

प्रारंभ में, समिति ने सभी सदस्यों और संभावित बोलीदाताओं के प्रतिनिधियों का स्वागत किया और सामान्य तौर पर निविदा के दायरे की जानकारी दी और उसके बाद सहायक कुलसचिव ( भंडार एवं क्रय) से बोली लगाने वालों को निविदा की मुख्य विशेषताओं के बारे में जानकारी देने का अनुरोध किया।

At the outset, the committee welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the tender and thereafter requested Assistant Registrar (S&P) to brief the bidders on the salient features of the tender.

उपस्थित प्रतिनिधि दिए गए उत्तरों से संतुष्ट थे और यह सूचित किया गया था कि प्री बिड कॉन्फ्रेंस के दौरान की गई चर्चा के अनुसार दिए गए सुधार- / परिवर्धन / स्पष्टीकरण को IISER पुणे की वेबसाइट पर होस्ट किया जाएगा और सभी संभावित बोलीदाताओं को बोली दस्तावेजों में निर्धारित अनुसार अपनी बोली जमा करने से पहले प्री-बिड सम्मेलन की कार्यवाही का संज्ञान लेना आवश्यक है।

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

हमारी आईआईएसईआर वेबसाइट [www.iiserpune.ac.in](http://www.iiserpune.ac.in) और GeM portal पर जारी नोटिस के अन्य नियम और शर्तें अपरिवर्तित रहेंगी। इस संबंध में और कोई पत्राचार नहीं किया जाएगा।

The other terms & conditions of the notice issued on our IISER website [www.iiserpune.ac.in](http://www.iiserpune.ac.in) and GeM portal will remain unchanged. No more correspondence in this regard will be entertained.

बैठक अध्यक्ष के धन्यवाद प्रस्ताव के साथ समाप्त हुई।  
The meeting ended with vote of thanks to the Chair.

07/03/2025

**Sd/-**  
सहायक कुलसचिव (भंडारण एवं क्रय)  
Sr. Assistant Registrar (S&P)

TECHNICAL AND COMMERCIAL QUERIES AND CLARIFICATION**PRE-BID CONFERENCE FOR ITEM DESCRIPTION - PROCUREMENT OF FOUR PORTS GLOVE BOX**

S. No.	Query/Clarification Sought	Clarification / Amendment
1	<p><b><u>Pt No. 5 : Gas purification</u></b>  Purifier should be customised and located underneath the glove box - Is it required?  Should be compatible with voice assistance when upgraded - please confirm if its for future upgrade or current requirement.</p>	<p>Yes, it is required.  Yes, this is current requirement.  Tender specifications prevails</p>
2	<p><b><u>Pt No. 11 : Solvent adsorption unit:</u></b>  Solvent adsorption unit with suitable adsorbent at least 5kgs  Marks allotted to this is only 2 points we request you to please change the same and increase the marks as it is an important factor in the glove box purification.</p>	<p>Solvent adsorption unit is a very normal thing and thus the allotted marks is justified.  Tender specifications prevails</p>
3	<p><b><u>Pt No. 13: Control System</u></b>  PLC Controlled with Color Touch panel for operation of all Glove box functions, including graphical trend of box pressure, oxygen and moisture levels. Provision for remote monitoring of glove box parameters. Remote monitoring of glove box parameters including sending alerts and maintenance notifications and tutorials for servicing and application easily downloadable.  Reminders for spare replacement.</p> <p>Upgrade Possibility for voice assistance upgrade is required. - please clarify if it is to be quoted in current requirement or it is for future upgrades</p>	<p>This is current requirement.  PLC control panel is one of the most important parts of the entire system and thus the allotted marks are fully justified.  Tender specifications prevails</p>

	<p>Marks allotted to this is 15 points we request you to please change the same since it's only a PLC control panel for glove box.</p>	
4	<p><b><u>Pt No. 13: Other Essential Requirements:</u></b>  Spare glovex - 4 pairs  Spare O ring for gloves - 4 Pairs  Spare O ring for miniantchamber - 4 pairs.  Spare package of activated charcoal - 6 full loads.  04 Numbers of Argon Gas Cylinders with regulators  02 Numbers of Calibration Gas Cylinders with regulators  10 KVA UPS with 30 mins backup</p> <p>Marks allotted to this is only 2 points we request you to please change the same and increase the marks as it is an important factor in running the glove box for more years without any problem</p>	<p>These are regular items for the system and thus the allotted marks are fully justified.</p> <p>Tender specifications prevails</p>
5	<p><b><u>Pt No. 13: Critical Terms &amp; Conditions for qualification:</u></b>  Vendors are required to provide ONLY relevant brochures  Upgrade features should be supported by specifications and pricing  Just mention of complied /meeting specification will not lead to technical qualification.  Warranty/Guarantee: Three Years  AMC: Three years after warranty period</p> <p>Marks allotted to this is only 2 points we request you to please change the same and increase the marks as it is an important factor in running the glove box for more years without any problem.</p>	<p>These are very common for the system and thus the allotted marks are fully justified.</p> <p>Tender specifications prevails</p>