

INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH

An Autonomous Institution, Ministry of Education, Govt. of India

Dr. Homi Bhabha Road, Pashan Pune – 411008.

Tel: +91-020-25898017; Email: purchase@iiser.pune.ac.in

Website: www.iiserpune.ac.in

DATE: 24/03/2025

TENDER REFERENCE NO: IISER/PUR/2307/24

Tender ID: 2025_IISRP_852694_1

CORRIGENDUM

With tender reference no IISER/PUR/2307/24 dated 10/3/2025, for procurement of Low-vacuum plasma cleaner for “High-throughput single cell analysis system”. It is clarified that tendered specifications are amended as below:

Tendered Specification	Amendment in Specification
Vacuum pressure (lowest point) 0.003 mbar Plasma system should be in working condition on delivery for testing.	Vacuum pressure (lowest point) Less than 0.05 mBar upto 0.003 mBar. Plasma system should be in working condition on delivery for testing.
Frequency 40 KHz	The plasma frequency to be minimally in KHz range, with MHz also acceptable.


Sr. Assistant Registrar(S&P)