



भारतीय विज्ञान शिक्षा एवं अनुसंधान संस्थान पुणे

INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH PUNE

IISER/PUR/1576/25

PREBID CLARIFICATION ON CPP TENDER NUMBER - IISER/PUR/1576/25

आइटम विवरण- अतिचालक चुंबक और सहायक उपकरण की खरीद

Item Description- Procurement of Superconducting Magnet and Accessories

सुपरकंडक्टिंग मैग्नेट और सहायक उपकरणों की खरीद के लिए संस्थान की वेबसाइट www.iiserpune.ac.in और सीपीपी पोर्टल पर 06/02/2026 को प्रकाशित खुली निविदा का संदर्भ लें।

Refer to an open tender published on the Institute website www.iiserpune.ac.in and on the CPP Portal on 06/02/2026 for procurement of Superconducting Magnet and Accessories.

प्री-बिड मीटिंग 13/02/2026 को शाम 3.00 बजे आयोजित की गई और बैठक का कार्यवृत्त निम्नानुसार है:

Pre-Bid meeting was held on 13/02/2026 at 3.00 PM and minutes of meeting is as under:

प्रारंभ में, समिति ने सभी सदस्यों और संभावित बोलीदाताओं के प्रतिनिधियों का स्वागत किया और सामान्य तौर पर निविदा के दायरे की जानकारी दी और उसके बाद उप कुलसचिव (भंडार एवं क्रय) से बोली लगाने वालों को निविदा की मुख्य विशेषताओं के बारे में जानकारी देने का अनुरोध किया।

At the outset, the committee welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the tender and thereafter requested Deputy Registrar (S&P) to brief the bidders on the salient features of the tender.

उपस्थित प्रतिनिधि दिए गए उत्तरों से संतुष्ट थे और यह सूचित किया गया था कि प्रीबिड कॉन्फ्रेंस के दौरान की गई चर्चा के अनुसार दिए गए सुधार - परिवर्धन / स्पष्टीकरण को /IISER पुणे की वेबसाइट पर होस्ट किया जाएगा और सभी संभावित बोलीदाताओं को बोली दस्तावेजों में निर्धारित अनुसार अपनी बोली जमा करने से पहले प्री-बिड सम्मेलन की कार्यवाही का संज्ञान लेना आवश्यक है।

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

हमारी आईआईएसईआर वेबसाइट www.iiserpune.ac.in और CPP Portal पर जारी नोटिस के अन्य नियम और शर्तें अपरिवर्तित रहेंगी। इस संबंध में और कोई पत्राचार नहीं किया जाएगा।

The other terms & conditions of the notice issued on our IISER website www.iiserpune.ac.in and CPP Portal will remain unchanged. No more correspondence in this regard will be entertained.

बैठक अध्यक्ष के धन्यवाद प्रस्ताव के साथ समाप्त हुई।

The meeting ended with vote of thanks to the Chair.

20/02/2026

Sd/-

उप कुलसचिव (भंडारण एवं क्रय)

Deputy Registrar (S&P)

TECHNICAL AND COMMERCIAL QUERIES AND CLARIFICATIONPRE-BID CONFERENCE FOR PROCUREMENT OF **SUPERCONDUCTING MAGNET AND ACCESSORIES**

S. No.	Query/Clarification Sought	Clarification / Amendment
1.	<p><u>CHAPTER-2: INSTRUCTIONS TO BIDDERS, CLAUSE NO. 2 DELIVERY PERIOD</u></p> <p>A. Further to the pre-bid meeting regarding your requirement of Superconducting Magnet and Accessories, we request you to kindly extend the delivery period to 9 months instead of 75 days. Due to customize design, these type of systems require longer delivery schedule.</p> <p>B. Considering the manufacturing and logistics cycle for superconducting magnets, please clarify whether the stipulated delivery period of 75 days is firm or can be finalized mutually based on OEM production schedules.</p>	<p><u>CHAPTER-2: INSTRUCTIONS TO BIDDERS, CLAUSE NO. 2 DELIVERY PERIOD</u></p> <p>The delivery period is amended from 75 days to 180 days from the date of establishment of the Letter of Credit.</p>
2.	<p>Kindly confirm whether the USD 210,000 ceiling is applicable strictly to the base magnet system and whether optional accessories quoted separately will be excluded from L1 price evaluation.</p>	<p>While the Institute has disclosed the overall estimated tender value, Institute do not declare item-wise internal budget allocations or individual cost ceilings. Bidders are advised to submit their most competitive bids for the complete system based on the technical specifications provided.</p> <p>The determination of the L1 (Lowest Evaluated Bidder) will be based solely on the price of the Base System and mandatory technical requirements. Prices for optional accessories/items must be quoted separately and will not be factored into the primary financial evaluation.</p>

<p>3.</p>	<p><u>CHAPTER 4- SCHEDULE OF REQUIREMENTS, SPECIFICATIONS & ALLIED TECHNICAL DETAILS</u></p> <p>We propose a warranty structure of three years in total – two years comprehensive followed by one year non-comprehensive. Additionally, since cryocooler servicing is typically treated as a consumable maintenance activity, kindly confirm whether periodic cryocooler service can be excluded from the warranty scope.</p>	<p><u>CHAPTER 4 - SCHEDULE OF REQUIREMENTS, SPECIFICATIONS & ALLIED TECHNICAL DETAILS</u></p> <p>Tender specification prevails. Cryocooler servicing also follows the tender specifications for warranty.</p>
<p>4.</p>	<p><u>CHAPTER 4- SCHEDULE OF REQUIREMENTS, SPECIFICATIONS & ALLIED TECHNICAL DETAILS</u></p> <p>It's a room temperature bore system and so there is a tube and the only access is the top and the bottom as there are no windows. So, how do you perform window-optical work(spectroscopy) if there are no windows and only access are the ends? Is the spectroscopy being done using fiber optics that come down to the sample or are you using vertical windows in the room temperature bore? If it has to have CRYOSTAT Windows, it's a split coil magnet, which is different than the traditional magnet. So, please clarify how is the spectroscopy being done in the room temperature bore?</p>	<p><u>CHAPTER 4- SCHEDULE OF REQUIREMENTS, SPECIFICATIONS & ALLIED TECHNICAL DETAILS</u></p> <p>Yes, this is a room temperature bore system, which means that the access to magnetic field is from top and bottom. We will insert our cryostat finger (separate system which we have) on the bore from the bottom. The sample will sit on the top of the finger, which will be covered with a window. The light will fall from the top on the sample, through window, and reflect back for spectroscopy.</p>
<p>5.</p>	<p>Possibility to quote Chiller in local currency in BoQ</p>	<p>Bidders may quote for the Chiller in Indian Rupees (INR)</p>