

भारतीय विज्ञान शिक्षा एवं अनुसंधान संस्थान पुणे
INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH

An Autonomous Institution, Ministry of Education, Govt. of India

Dr. Homi Bhabha Road, Pashan Pune – 411008.

Tel: +91-020-25898017; Email-purchase@iiser.pune.ac.in

Website: www.iiserpune.ac.in

DATE : 07/08/2026

निविदा संदर्भ संख्या / **TENDER REFERENCE NO: IISER/PUR/0463/26**

निविदा आईडी / **TENDER ID: 2026_DoHE_846273_1**

CORRIGENDUM

With Tender reference no. IISER/PUR/0463/26 dated 27/07/2026 for Linkam PE120-XY System. It is notified that critical dates of the tender are amended as below.

TECHNICAL CORRIGENDUM

Tendered / Pre-Bid Specification	Amendment in Specification
1. Linkam PE120-XY System; PE120-XY Heating/cooling Peltier Stage. Simple heating and cooling stage for microscope slides, with a temperature range from -20°C to 120°C. To fit XY tables with 160mm x 110mm recess or can be used on the top surface of XY tables without recess, with the optical axis at 81.9 mm from rear edge of the PE120 stage (74mm from the centres of the slide-holder holes). Comprising: PE120-XY Heating/cooling Peltier Stage, T96-P System Controller (without linkpad), Temperature control computer software to be installed.	1. Linkam PE120-XY System; PE120-XY Heating/cooling Peltier Stage. Simple heating and cooling stage for microscope slides, with a temperature range from -25°C to 120°C. To fit XY tables with 160mm x 110mm recess or can be used on the top surface of XY tables without recess, with the optical axis at 81.9 mm from rear edge of the PE120 stage (74mm from the centers of the slide-holder holes). Comprising: PE120- XY Heating/cooling Peltier Stage, T96-P System Controller with linkpad for temperature control.
3. Evident LUCPLFLN40X2 Semi plan apochromatic objective 40X/0.6 WD 2.70-4.0 (Correction collar). (Any other Evident objective with Air/Dry immersion, with higher NA, and longer WD (up to 8 mm) of 40x or 50x magnification, could also be quoted)	3. Evident LUCPLFLN40X2 Semi plan apochromatic objective 40X/0.6 WD 2.70-4.0 (Correction collar) OR (Evident Long Working Distance Objective of 50x/.050 WD of 10.6mm or higher and OFN 26.5mm Semi apochromat. The quoted objectives should be compatible with Olympus microscope and to Linkam Stage.

DATE CORREGENUDUM

Sr. No.	Particulars	Date	Time
1	Bid Submission Close Date	14/08/2026	15:00 Hrs
2	Opening of Bids	17/08/2026	14:00 Hrs

Sd/-

उप कुलसचिव (भंडारण एवं क्रय)
Deputy Registrar (S&P)