



भारतीय विज्ञान शिक्षा एवं अनुसंधान संस्थान पुणे
INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH PUNE

PREBID CLARIFICATION ON GeM TENDER NUMBER - GEM/2026/B/7845264
(IISER/PUR/1896/26)

आइटम विवरण- टाइम-डोमेन इलेक्ट्रोमैग्नेटिक (TEM) इंस्ट्रूमेंट की खरीद

Item Description- Procurement of Time-Domain Electromagnetic (TEM) Instrument

टाइम-डोमेन इलेक्ट्रोमैग्नेटिक (TEM) इंस्ट्रूमेंट की खरीद के लिए, इंस्टीट्यूट की वेबसाइट www.iiserpune.ac.in और GeM पोर्टल पर 29/07/2026 को प्रकाशित ओपन टेंडर देखें।

Refer to an open tender published on the Institute website www.iiserpune.ac.in and on the GeM Portal on 29/07/2026 for procurement of Time-Domain Electromagnetic (TEM) Instrument.

प्री-बिड मीटिंग 04/08/2026 को शाम 3.00 बजे आयोजित की गई और बैठक का कार्यवृत्त निम्नानुसार है:

Pre-Bid meeting was held on 04/08/2026 at 3.00 PM and minutes of meeting is as under:

शुरुआत में, समिति ने सभी सदस्यों और संभावित बोलीदाताओं के प्रतिनिधियों का स्वागत किया और उन्हें टेंडर के दायरे के बारे में सामान्य जानकारी दी; इसके बाद, सेक्शन ऑफिसर (S&P) से कहा गया कि वे बोलीदाताओं को टेंडर की मुख्य विशेषताओं के बारे में जानकारी दें।

At the outset, the committee welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the tender and thereafter requested Section Officer (S&P) to brief the bidders on the salient features of the tender.

उपस्थित प्रतिनिधि दिए गए उत्तरों से संतुष्ट थे और यह सूचित किया गया था कि प्री बिड कॉन्फ्रेंस के दौरान की गई चर्चा के अनुसार दिए गए सुधार- / परिवर्धन / स्पष्टीकरण को IISER पुणे की वेबसाइट पर होस्ट किया जाएगा और सभी संभावित बोलीदाताओं को बोली दस्तावेजों में निर्धारित अनुसार अपनी बोली जमा करने से पहले प्री-बिड सम्मेलन की कार्यवाही का संज्ञान लेना आवश्यक है।

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

हमारी आईआईएसईआर वेबसाइट www.iiserpune.ac.in और GeM portal पर जारी नोटिस के अन्य नियम और शर्तें अपरिवर्तित रहेंगी। इस संबंध में और कोई पत्राचार नहीं किया जाएगा।

The other terms & conditions of the notice issued on our IISER website www.iiserpune.ac.in and GeM portal will remain unchanged. No more correspondence in this regard will be entertained.

बैठक अध्यक्ष के धन्यवाद प्रस्ताव के साथ समाप्त हुई।

The meeting ended with vote of thanks to the Chair.

07/08/2026

Sd/-
उप-कुलसचिव (भंडारण एवं क्रय)
Deputy Registrar (S&P)

TECHNICAL AND COMMERCIAL QUERIES AND CLARIFICATION**PRE-BID CONFERENCE FOR ITEM DESCRIPTION- PROCUREMENT OF TIME-DOMAIN ELECTROMAGNETIC (TEM) INSTRUMENT**

S. No.	Query/Clarification Sought	Clarification / Amendment
1	Transmitter specifications: Bidder requested to amend repetition frequency: For LM 2.5 kHz or more For HM 1.5 j Hz or more	Transmitter specifications: As mentioned in the general description of the instrument, the system should support acquisition of good-quality data while walking, which requires a high repetition frequency to allow sufficient stacking. Therefore, the instrument should have a repetition frequency of greater than 1.0 kHz.
2	Receiver specifications: Bidder requested to specify the first gate specifications: LM 5 μ s or less HM 10 μ s or less	Receiver specifications: Instead of specifying the first-gate time separately, the specification defines the time to the center of the first gate (Sr. No. 18), which adequately addresses the required first-gate timing. Therefore, no separate specification for the first-gate time is required.
3	Bidder requested to confirm the communication between Transmitter and Receiver through WiFi.	Yes. Communication between the transmitter and receiver should be through Wi-Fi to facilitate continuous and efficient data recording while the user is walking.
4	Bidder requested to confirm the Understood data Collection is in continuous measurements using the transmitter and receiver.	Yes. As specified in the tender, the system shall be capable of recording data continuously while the user is walking. This implies continuous data acquisition during operation.