



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH PUNE

CLARIFICATION ON TENDER NUMBER - IISER-PUR-1157-20

ITEM DESCRIPTION- PROCUREMENT OF RESEARCH GRADE STEADY STATE FLUORESCENCE SPECTROMETER.

Refer IISER Pune open tender number IISER-PUR-1157-20 for procurement of Research Grade Steady State Fluorescence Spectrometer.

Pre-Bid meeting was held on May 4th, 2021 at 4.00 PM via video conferencing and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website www.iiserpune.ac.in will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

4.5.2021

Sd/-
Assistant Registrar (S&P)



IISER PUNE

PROCUREMENT OF RESEARCH GRADE STEADY STATE FLUORESCENCE SPECTROMETER

TECHNICAL QUERIES AND CLARIFICATION

S.No	Query/Clarification Sought	Clarification / Amendment
1	<p>1. You have asked Solid Sample Holder for thin films, powders, pellets, paper, fibers, or microscopic slides measurements Reply: We have not seen any band pass filters for the highly scattering solid samples such as films/powders. Normally 320nm and 370nm bandpass filters are recommended. Please confirm about the band pass filters.</p> <p>2. You have asked PC with specific requirements and Apple Mac Book Air Reply: We are not selling PC. However, we can include the quote from PC-Apple MAC supplier which will be in Indian rupees. Please note there is no provision in the BOQ for entering the INR price.</p>	<p>Band pass filters will be purchased later, when required. May add as accessories. No Amendment of technical specification is required.</p> <p>To include PC please attach the quote in PDF along with the price schedule of BOQ. While determining the L1 the total price in BOQ and pdf will be taken into consideration</p>
2	<p>1. Our instrument is having 4 turret FOV lens holder as standard. You have opted for 5 turret, can you specify as 4-5 turret</p> <p>2. Diameter of an atom falls in the range of 0.1 -0.5nm, you have specified measurement range and repeatability as 0.1nm & 0.01nm, we assume the numbers are theoretically better numbers. We request you to specify as resolution 1nm, repeatability 0.1nm so that specifications are more realistic.</p> <p>3. Since many of your optional features are standard in our instrument can you increase the budget by Rs 5 lakhs</p>	<p>Query not relevant to this tender enquiry</p>