



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH
PUNE

CLARIFICATION ON TENDER NUMBER - IISER-PUR-0231-16

ITEM DESCRIPTION- PROCUREMENT OF CAGE CHANGING STATION

Refer our Press Tender Notice No.IISER/S&P/05/2016 dated 05.8.2016 for procurement of Cage Changing Station Tender Reference Number - IISER-PUR-0231-16.

Pre-Bid meeting was held on Aug 12th , 2016 at 2.00 PM and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website [www.iiserpune.ac .in](http://www.iiserpune.ac.in) will remain unchanged.No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

12.8.2016

Sd/-
Assistant Registrar (S&P)



IISER PUNE

PRE-BID CONFERENCE FOR PROCUREMENT OF CAGE CHANGING STATION**TECHNICAL QUERIES AND CLARIFICATION**

TENDER NUMBER - IISER-PUR-0231-16

DATE : 12.8.16

S.No	Query/Clarification Sought	Clarification / Amendment
1	CHAPTER 4 - SCHEDULE OF REQUIREMENTS, SPECIFICATIONS & ALLIED TECHNICAL DETAILS. Point No. 4. Functional data can be recorded and saved. What data you want to record and save?	CHAPTER 4 - SCHEDULE OF REQUIREMENTS, SPECIFICATIONS & ALLIED TECHNICAL DETAILS. Point No. 4 - Functional data can be recorded and saved. is deleted from the specification



IISER PUNE

PRE-BID CONFERENCE FOR PROCUREMENT OF CAGE CHANGING STATION

COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-0231-16

DATE : 12.8.16

S.No	Query/Clarification Sought	Clarification / Amendment
1	<p>CHAPTER-2 : INSTRUCTIONS TO BIDDERS Clause No. 2. Delivery Period / Timeliness. 2.1: The deliveries & installation must be completed within 02 months after placement of Purchase order</p> <p>It is not possible to supply and install the machine within 2 months after the receipt of PO. Because 10-12 weeks are required to manufacture 8 Cage Change Stations. After that 4-5 weeks are required for the shipping from Italy to India. So practically it is not possible to supply the material within 2 months. Installation is not a problem. We can install the machines within 2-3 days. Request to change the delivery period /timeliness.</p>	<p>CHAPTER-2 : INSTRUCTIONS TO BIDDERS Clause No. 2. Delivery Period / Timeliness is amended to :</p> <p>The deliveries & installation must be completed within 03 months after placement of Purchase order</p>