



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH  
PUNE

CLARIFICATION ON TENDER NUMBER - IISER-PUR-0254-15

ITEM DESCRIPTION- PROCUREMENT OF TI:SAPPHIRE, TUNABLE, FEMTOSECOND LASER

Refer our Press Tender Notice No.IISER/S&P/04/15 dated 25.6.2015 for procurement of TI:SAPPHIRE, TUNABLE, FEMTOSECOND LASER . Tender Reference Number - IISER-PUR-0254-15.

Pre-Bid meeting was held on July 03<sup>rd</sup> , 2015 at 3.00 PM and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website [www.iiserpune.ac.in](http://www.iiserpune.ac.in) will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

3.7.2015

Sd/-  
Assistant Registrar (S&P)



IISER PUNE

**PRE-BID CONFERENCE FOR PROCUREMENT OF TI:SAPPHIRE, TUNABLE, FEMTOSECOND LASER**

**TECHNICAL QUERIES AND CLARIFICATION**

TENDER NUMBER - IISER-PUR-0254-15

DATE : 3.7.15

S.No	Query/Clarification Sought	Clarification / Amendment
1	BBO crystal thickness (accessory)	around 2 mm thickness
2	Optical Delay line (accessory)	up to 2 to 3 ns
3	whether the required tuning range of 690-1040 nm must be met or whether $800 \pm 20$ nm will be acceptable	Tuning range of around 690-1040 nm must be met. $800 \pm 20$ is not acceptable



IISER PUNE

**PRE-BID CONFERENCE FOR PROCUREMENT OF TI:SAPPHIRE, TUNABLE, FEMTOSECOND LASER**

## COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-0254-15

DATE : 3.7.15

S.No	Query/Clarification Sought	Clarification / Amendment
	-----NIL-----	-----NIL-----