



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH
PUNE

CLARIFICATION ON TENDER NUMBER - IISER-PUR-1207-14

ITEM DESCRIPTION- PROCUREMENT OF 4-PORT GLOVE BOX WITH INTEGRATED SOLVENT PURIFICATION SYSTEM

Refer our Press Tender Notice No.IISER/S&P/10/14 dated 10.12.2014 for procurement of 4-Port Glove Box with Integrated Solvent Purification System. Tender Reference Number - IISER-PUR-1207-14.

Pre-Bid meeting was held on December 23rd, 2014 at 2.30 PM and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website www.iiserpune.ac.in will remain unchanged.No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

23.12.14

Sd/-
Assistant Registrar (S&P)



IISER PUNE

PRE-BID CONFERENCE FOR PROCUREMENT OF 4-PORT GLOVE BOX WITH INTEGRATED SOLVENT PURIFICATION SYSTEM

TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-1207-14

DATE :23.12.14

S.No	Query/Clarification Sought	Clarification / Amendment
1	<p>Page 24, Point Number E3:</p> <p>Does the glove box require double purification columns?</p>	<p>Page 24, Point Number E3:</p> <p>We have mentioned re-generable single filter gas purifier in our specifications. Nonetheless, if the option for double purification column is available, bidders may quote it separately.</p> <p>There will be no change in specifications.</p>
2	<p>Page 26, Point Number B2 and B4:</p> <p>Does solvent purification system require refill material?</p>	<p>Page 26, Point Number B2 and B4:</p> <p>We have mentioned in our specifications that the double column filters in Solvent Purification System need to be activated prior to shipment. The entire filter columns will be replaced after they are saturated. Since no activation will be done by us, we will not require refill material.</p>

3	<p>Page 25, Point Number H:</p> <p>Does integration with spectrometer/electrochemical workstation or with any external devices required?</p>	<p>Page 25, Point Number H:</p> <p>Yes, the glove box will have spectrometer/electrochemical workstation inside it. The required feedthroughs for external devices have been mentioned in the specifications. If there are additional options, bidders may quote it separately.</p>
4	<p>Page 25, Point Number I</p> <p>Is it mandatory to have a solid-state sensor for oxygen and moisture in the glove box?</p>	<p>Page 25, Point Number I:</p> <p>Yes, it is absolutely mandatory to have solid-state oxygen and moisture sensor only for the glove box.</p>
5	<p>Page 26:</p> <p>Both Glove Box and Solvent Purification should be from single manufacturer?</p>	<p>Page 26:</p> <p>Yes, we have clearly mentioned in our specifications that both the Glove Box and Solvent Purification should be from single manufacturer only.</p>



IISER PUNE

**PRE-BID CONFERENCE FOR PROCUREMENT OF 4-PORT GLOVE BOX WITH INTEGRATED
SOLVENT PURIFICATION SYSTEM**

COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-1207-14

DATE :23.12.14

S.No	Query/Clarification Sought	Clarification / Amendment
	-----NIL-----	-----NIL-----