



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH
TIRUPATI

CLARIFICATION ON TENDER NUMBER - IISERT-PUR-0122-16

ITEM DESCRIPTION- TABLE TOP NMR

Refer our Press Tender Notice No.IISER/S&P/09/16 dated 7.10.2016 for procurement of Table Top NMR. Tender Reference Number - IISERT-PUR-0122-16

Pre-Bid meeting was held on October 20th , 2016 at 14.00 and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additons / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website [www.iiserpune.ac .in](http://www.iiserpune.ac.in) will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

20.10.2016

Sd/-
Assistant Registrar (S&P)

IISER TIRUPATI

PRE-BID CONFERENCE FOR PROCUREMENT OF TABLE TOP NMR

TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER – IISERT-PUR-0122-16

DATE : 24.10.16

S.No	Query/Clarification Sought	Clarification / Amendment
01	<p>Page No - 22, Chapter 4, Point no. 9</p> <p>Sensitivity: should be > 100:1 for 1% ethylbenzene with single scan.</p> <p>Suggestion by one of the bidders: We offer sensitivity 40: 1 for 1% ethylbenzene with single scan. Since we offer 60 MHz , 1.4 T NMR ,This sensitivity is sufficient to get data for 1H/19F</p>	<p>More sensitivity is desired.</p> <p>Tender specification prevails.</p>
02	<p>Page No -22, Chapter 4, Point no. 10</p> <p>Resolution at 50% should be < 0.5Hz.</p> <p>Suggestion by one of the bidders “We offer resolution at 50% <1.2 Hz. Since we offer 60 MHz, 1.4 T NMR, this resolution is sufficient to get data for 1H/19F”</p>	<p>Higher resolution is desired.</p> <p>Tender specification prevails.</p>



IISER TIRUPATI

PRE-BID CONFERENCE FOR PROCUREMENT OF TABLE TOP NMR

COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER – IISERT-PUR-0122-16

DATE : 24.10.16

S.No	Query/Clarification Sought	Clarification / Amendment
	-----NIL-----	-----NIL-----