



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH  
PUNE

CLARIFICATION ON TENDER NUMBER - IISER-PUR-1329-19

ITEM DESCRIPTION- PROCUREMENT OF CNS INSTRUMENT

Refer IISER Pune open tender number IISER-PUR-1329-19 dated 9.3.2020 for procurement of CNS Instruments.

Pre-Bid meeting was held on May 11<sup>th</sup>, 2020 at 2.30 PM via video conferencing and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website [www.iiserpune.ac .in](http://www.iiserpune.ac.in) will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

11.5.2020

Sd/-  
Assistant Registrar (S&P)



## IISER PUNE

**PRE-BID CONFERENCE FOR PROCUREMENT OF CNS INSTRUMENT**

## TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-1329-19

DATE : 11.5.2020

S.No	Query/Clarification Sought	Clarification / Amendment
1	Calibration can be provided on chargeable basis for following parameters: Temperature, Shock(CURRENT), Distance travelled, Calibration for the mazes. As far as validation is concerned it is not applicable for this equipment.	The parameters indicated are fine. However, for the period of 3 years from purchase and installation, annual calibration should be provided without any charge.
2	IQ/OQ will be submitted in the form of SOP's as already discussed with the user. User training can be done as and when demanded as these are research based instruments only.	OK

3	Tail Flick: Please mention just “HEAT SOURCE” instead IR Heat Source ?? whereas our system works on the principal of visible light to prevent unnecessary conditioning of animals.	OK. Other indicated technical specifications are unchanged.
4	Hot & Cold: -10deg ??? usually starts from Zero degrees to 70 degrees and not -10.	0 to 70 degrees C is OK
5	ROTAROD: You have mentioned 80RPM in your tender document but 0 TO 45 RPM range encompass all experimental needs and we offer up to 45 RPM.	No change in tender specifications. The bidder may quote system with lower speed as optional.
6	Video Tracking: We need to understand the objective behind using video camera for active and passive protocols.	Main quote should have the camera for the mazes. A separate camera for the shuttle box should be quoted as an optional.
7	Shuttle Box: The term “OPTIONAL ACCESSORIES” is to be removed since the experimental needs demand active and passive avoidance and the sliding door is the integral part of the equipment.	OK. Sliding door should be quoted as a part of the main quote. The system quoted should be capable of active and passive avoidance tests.



IISER PUNE

## PRE-BID CONFERENCE FOR PROCUREMENT OF CNS INSTRUMENT

### COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-1329-19

DATE : 11.5.2020

S. No	Query/Clarification Sought	Clarification / Amendment
1.	IISER Pune has asked for Solvency Certificate which no bank provides hence it should not be made mandatory?	Please check with your bank. Banks provides solvency certificate. Hence, you may submit solvency certificate not older than 12 months issued by scheduled/nationalized bank with which your company holds current account.
2.	We will be quoting in INR and can provide the MSME Certificate for exemption of EMD ? Is it O.K. ?	Micro and Small Enterprises (MSE) must, along with their offer, provide proof of their being registered as MSE (indicating the terminal validity date of their registration) for the item tendered. However EMD and tender fees is exempted for procurement of only goods produced and services rendered by MSEs and not for any trading activities by them.

3.	Given the situation of COVID-19, there are situations that shipment may get delayed on account of larger workloads at everybody's end hence we feel that 90 days delivery period should be allowed instead of 45 days as mentioned in the tender document. Will there be a force majeure clause from your side? This is to be confirmed by IISER-P.	Delivery period is extended from 45 days to 90 days.  Force majeure clause- Please refer page no 18 para 19 of tender document.
4.	Security Deposit/ PBG: IISER-P will extend the SD as PBG. Can we give Security Deposit for our Standard Warranty period only. I mean we will offer 1 Year STD warranty plus 2 Years Extended Warranty period to meet the requirement of 3 Years Warranty. Please check if SD can be given for 1 Year STD Warranty only ?	No.  Security Deposit/PBG should be submitted for entire warranty period of three years.
5.	IISER Pune have mentioned DSIR certificate will be provided. Nothing is mentioned about GSTCC. Please confirm since we will be quoting in INR i.e., we will be importing the goods from our principals and provide the goods on F.O.R basis to IISERP.	Yes, IISER Pune will provide GST concessional certificate for goods quoted in INR as they are required for research purpose.
6.	AMC: Everywhere in tender, you have asked to quote for AMC only but format given in Tender is for both AMC & CMC. Please check which one to quote as you will consider this for Evaluation purpose? will you take AMC or CMC values into account for arriving at L1 ??	Please quote for both Comprehensive AMC and Non - Comprehensive charges for AMC as per format given in Annexure H page no 38 along with price schedule in pdf format.  For evaluation purpose AMC value will not be taken into account for arriving L1.
7.	Payment Terms are not flexible at all. We suggest for 100% PAYMENT (against shipment), as we are submitting 10% SD ?	No Change in payment terms
8.	<b>Provision for making ONLINE PAYMENT for tender fees, EMD, or any other fees, incidentals to be provided due to lockdown.</b>	Yes, ONLINE PAYMENT for tender fee and EMD amount can be deposited in IISER PUNE Bank account through net banking as mentioned below.

		<b>Name-IISER PUNE</b> <b>Bank-State Bank of India</b> <b>Branch-NCL Campus Branch, PUNE 411008</b> <b>Current A/c No. 30042605732</b> <b>IFSC-SBIN0003552</b>
--	--	--