



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH
TIRUPATI

CLARIFICATION ON TENDER NUMBER - IISERT-PUR-0019-16

ITEM DESCRIPTION- HIGH SPEED CENTRIFUGE

Refer our Press Tender Notice No.IISERT/S&P/01/16 dated 25.5.2016 for procurement of High Speed centrifuge. Tender Reference Number - IISERT-PUR-0019-16.

Pre-Bid meeting was held on June 02nd, 2016 at 14.30 and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website [www.iiserpune.ac .in](http://www.iiserpune.ac.in) will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

03.6.2016

Sd/-
Assistant Registrar (S&P)

IISER TIRUPATI

PRE-BID CONFERENCE FOR PROCUREMENT OF HIGH SPEED CENTRIFUGE

TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-0019-16

DATE : 03.06.16

S.No	Query/Clarification Sought	Clarification / Amendment
1	The Basic Centrifuge Speed Range is to be fixed from 22000 rpm - 26000 rpm instead of 26000 RPM only.	No change in specifications.
2	Change of Speed Control Accuracy to at least ± 25 rpm.	No change in specifications.
3	Rotor 3: Max Speed to 7,000 rpm	Amended specifications are Rotor 3- 7,000 RPM or more.
4	Rotor Material description in Each Rotor Asked to Aluminium (or) Carbon Fiber (or) Titanium.	No change in specifications.

Dr. Girish Ratanaparakhi

Dr. Sivakumar

Dr. Richa Rikhy

Dr. Sutirth Dey

Mr. Salim Shaikh



ANNEXURE -III

IISER TIRUPATI

PRE-BID CONFERENCE FOR PROCUREMENT OF HIGH SPEED CENTRIFUGE

TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-0019-16

DATE : 03.06.16

S.No	Query/Clarification Sought	Clarification / Amendment
01	The delivery Schedule to be extended for 3 Months instead of 2 Months .	Amended to 3 months

Dr. Girish Ratanaparakhi

Dr. Sivakumar

Dr. Richa Rikhy

Dr. Sutirth Dey

Mr. Salim Shaikh