

भारतीय विज्ञान शिक्षा एवं अनुसंधान संस्थान तिरुपती  
INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH TIRUPATI

Transit campus: Sree Rama Engineering College Campus, Rami Reddy Nagar, Karakambadi Road, Mangalam  
(B.O), Tirupati - 517 507



MENTORED BY  
Indian Institute of Science Education and Research Pune  
Dr. Homi Bhabha Raod, Pune : 411 008 (Maharashtra)  
[www.iiserpune.ac.in](http://www.iiserpune.ac.in)

INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH TIRUPATI

CLARIFICATION ON TENDER NUMBER - IISER-PUR-0125-15

ITEM DESCRIPTION- PROCUREMENT OF ANALYTICAL HPLC

Refer our Press Tender Notice No.IISER/S&P/08/15-16 dated 24.11.2015 for procurement of Analytical HPLC System. Tender Reference Number - IISER-PUR-0125-15.

Pre-Bid meeting was held on December 04<sup>th</sup> , 2015 at 2.00 PM and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additons / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website [www.iiserpune.ac.in](http://www.iiserpune.ac.in) will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

4.11.2015

Sd/-  
Assistant Registrar (S&P)



**IISER PUNE**  
**PRE-BID CONFERENCE FOR PROCUREMENT OF ANALYTICAL HPLC SYSTEM**

**TECHNICAL QUERIES AND CLARIFICATION**

TENDER NUMBER - IISERT-PUR-0125-15

DATE : 4.12.15

S.No	Query/Clarification Sought	Clarification / Amendment
1	<p>Binary/Quaternary Gradient delivery system with built-in on line degasser. Page No -25, Chapter 4, Point No. 1</p> <p>Clarification regarding whether bidders should quote Binary system or Quaternary Delivery System</p>	<p>Amended Specifications: Page No -25, Chapter 4, Point No. 1</p> <p><b>Binary High Pressure two separate Pump, Binary Gradient System</b></p> <p>Bidders are requested to quote their binary systems only.</p>
2	<p>Operating pressure: 6000 psi in all flow ranges.</p> <p>Page No -25, Chapter 4, Point No. 1 (First Clause)</p> <p>Suggestion by one of the bidder "Pump pressure drops if the flow rate is above 5 ml/minute. Hence we request you to change this point as " Pressure range 6 000psi upto 5ml flow rate"</p>	<p>We need above 5500 psi in all flow ranges (up to 10 ml/minute) for our analysis. Lowering this range to 5 ml/minute is not possible.</p>



IISER PUNE

**PRE-BID CONFERENCE FOR PROCUREMENT OF ANALYTICAL HPLC SYSTEM****COMMERCIAL QUERIES AND CLARIFICATION**

TENDER NUMBER - IISERT-PUR-0125-15

DATE : 4.12.15

S.No	Query/Clarification Sought	Clarification / Amendment
	-----NIL-----	-----NIL-----