



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH

TIRUPATI

CLARIFICATION ON TENDER NUMBER - IISERT-PUR-0202-15

ITEM DESCRIPTION- PROCUREMENT OF PARTICLE SIZE ANALYZER WITH ZETA POTENTIAL

Refer our Press Tender Notice No. IISERT/S&P/01/16 dated 25.5.2016 for procurement of Procurement of Particle Size Analyzer with Zeta Potential. Tender Reference Number - IISERT-PUR-0202-15.

Pre-Bid meeting was held on June 02nd, 2016 at 14.30 and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website www.iiserpune.ac.in will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

13.6.2016

Sd/-
Assistant Registrar (S&P)

IISER TIRUPATI

PRE-BID CONFERENCE FOR PROCUREMENT OF BENCH PARTICLE SIZE ANALYZER WITH ZETA POTENTIAL

TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISERT-PUR-0202-15

DATE: 13.6.16

S.No	Query/Clarification Sought	Clarification / Amendment
1	<p>Page No – 22, Chapter 4,</p> <p><i>“Particle Size measuring range: Minimum sample volume: 10-12µL”</i></p> <p>Request by one of the bidders to change the minimum sample volume to “up to 20 µL.”</p>	<p>Samples like proteins or solutions of nanostructured materials are available only in very minute volumes and hence the minimum volume of 10 µL is important.</p> <p>Tender Specification prevails.</p> <p>No change in the specification.</p>
2	<p>Page No – 22, Chapter 4,</p> <p><i>“Particle Size measuring range:</i></p> <p><i>Angles: 173° and 13 °”</i></p> <p>Request by one of the bidders to change the angles up to 175 ° and up to 15 °.</p>	<p>The amended specifications are</p> <p><i>“Particle Size measuring range:</i></p> <p><i>Angles: up to 175° and up to 15 °”</i></p>

3	<p>Page No – 22, Chapter 4,</p> <p>“Zeta potential and Protein Mobility: Minimum sample volume: 150 μL.”</p> <p>Request by one of the bidders to change the minimum sample volume to 350 μL.</p>	<p>Samples like proteins or solutions of nanostructured materials are available only in very small volumes and hence the minimum volume of 200 μL is important.</p> <p>The amended specifications are</p> <p>“Under Zeta potential and Protein Mobility: Minimum sample volume: 200 μL (20μL using diffusion barrier)</p>
4	<p>Page No – 23, Chapter 4,</p> <p>“Molecular Weight: Minimum sample volume: 12μL”</p> <p>Request by one of the bidders to change the minimum sample volume to 20 μL.</p>	<p>Samples like proteins or solutions of nanostructured materials are available only in very small volumes and hence the minimum volume of 15 μL is important.</p> <p>The amended specifications are</p> <p>“Under Molecular Weight: Minimum sample volume: 15 μL.”</p>
5	<p>Page No – 23, Chapter 4,</p> <p>“Molecular weight: Measurement range: 980Da – 2X10⁷ Da”</p> <p>Request by one of the bidders to change the measurement range to 1000 Da -2X10⁷ Da.</p>	<p>The amended specifications are</p> <p>“Molecular weight: Measurement range: 1000Da – 2X10⁷ Da”</p>
6	<p>Page No – 23, Chapter 4,</p> <p>“Light source: He-Ne laser 633nm, Max 10mW”</p> <p>Suggestion by one of the bidders that Diode lasers as Light Source is efficient.</p>	<p>The amended specifications are</p> <p>“Light Source: He-Ne laser 633 nm or Diode laser 658 nm”</p>



ANNEXURE -III

IISER TIRUPATI

PRE-BID CONFERENCE FOR PROCUREMENT OF BENCH PARTICLE SIZE ANALYZER WITH ZETA POTENTIAL

COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISERT-PUR-0202-15

DATE: 13.06.16

S.No	Query/Clarification Sought	Clarification / Amendment
	-----NIL-----	-----NIL-----