



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH
PUNE

CLARIFICATION ON TENDER NUMBER - IISERT-PUR-0161-15

ITEM DESCRIPTION- PROCUREMENT OF NITROGEN, HYDROGEN & ZERO AIR GENERATOR AND AIR COMPRESSOR

Refer our Press Tender Notice No.IISER/S&P/16/15-16 dated 28.1.2016 for procurement of NITROGEN, HYDROGEN & ZERO AIR GENERATOR AND AIR COMPRESSOR. Tender Reference Number - IISERT-PUR-0161-15.

Pre-Bid meeting was held on February 09th , 2016 at 2.00 PM and minutes of meeting is as under .

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website [www.iiserpune.ac .in](http://www.iiserpune.ac.in) will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

9.2.2016

Sd/-
Assistant Registrar (S&P)



IISER PUNE

**PRE-BID CONFERENCE FOR PROCUREMENT OF NITROGEN, HYDROGEN & ZERO AIR
GENERATOR AND AIR COMPRESSOR**

TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISERT-PUR-161-15

DATE : 9.2.16

S.No	Query/Clarification Sought	Clarification / Amendment
1	Performance Benchmarks - Page No - 26, Chapter 4, Clause - Zero air Generator Specification mentioned: particle < 0.01 / μm .	Specification is correct. Tender specification prevails.
2	Page No -26, Chapter 4 Same foot print as the N ₂ , H ₂ and Zero Air generator suitable for variety of gas generator combinations (in other words it should be stackable to minimize precious laboratory space)	All N ₂ , H ₂ and Zero Air should be stackable. Tender Specification prevails.

3	Page No - 26 , Chapter 4 Environment friendly for our Tropical Environment, as our country has 3 climatic conditions.	The room temperature is maintained at 25 °C and the same should be considered in our country as well. Tender specification prevails.
---	--	---



IISER PUNE

**PRE-BID CONFERENCE FOR PROCUREMENT OF NITROGEN, HYDROGEN & ZERO AIR
GENERATOR AND AIR COMPRESSOR**

COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISERT-PUR-161-15

DATE : 9.2.16

S.No	Query/Clarification Sought	Clarification / Amendment
	-----NIL-----	-----NIL-----