



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH
TIRUPATI

CLARIFICATION ON TENDER NUMBER - IISERT-PUR-0147-16

ITEM DESCRIPTION- AUTOMATED PURIFICATION SYSTEM (FPLC) FOR BIOMOLECULES

Refer our Press Tender Notice No.IISER/S&P/09/16 dated 7.10.2016 for procurement of Automated Purification System (FPLC) For Biomolecules. Tender Reference Number - IISERT-PUR-0147-16

Pre-Bid meeting was held on October 20th , 2016 at 14.00 and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website [www.iiserpune.ac .in](http://www.iiserpune.ac.in) will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

20.10.2016

Sd/-
Assistant Registrar (S&P)



IISER TIRUPATI
PRE-BID CONFERENCE FOR PROCUREMENT OF AUTOMATED PURIFICATION SYSTEM
(FPLC) FOR BIOMOLECULES

TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER – IISERT-PUR-0147-16

DATE : 24.10.16

S.No	Query/Clarification Sought	Clarification / Amendment
01	Can the minimum volume of the special injector loop be specified at 90 ml	The FPLC should be equipped with an injector valve through which the operator can inject samples in the volume range of 0.1 ml to at least 90 ml (using a special injector loop).
02	Can the wavelength range and numbers be specified in case of the multi-wavelength detector	The FPLC system should be equipped with an inline multiple wavelength detector capable of monitoring at least 3 wavelengths simultaneously in the range from of 190 nm- 900 nm.
03	Should the FPLC require any buffer inlet valves	The FPLC should be equipped with two buffer inlet valves, one each connected to the two available sample pumps.
04	Should the FPLC be equipped with a gradient mixer.	The FPLC be equipped with a gradient mixer with magnetic stirrer which can operate at the above flow rate and mix 0-100% of buffer B with the corresponding amount of buffer A.



IISER TIRUPATI

PRE-BID CONFERENCE FOR PROCUREMENT OF AUTOMATED PURIFICATION SYSTEM (FPLC) FOR BIOMOLECULES

COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER – IISERT-PUR-0147-16

DATE : 24.10.16

S.No	Query/Clarification Sought	Clarification / Amendment
	-----NIL-----	-----NIL-----