



IISER PUNE

TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR- 0523-16

Date: 14-10-2016

S.No	Query/Clarification Sought	Clarification / Amendment
1	Specifications: 1, c. Pulse Energy = 750 mJ or higher d. Pulse repetition rate = 1 - 10 Hz or more Merging these two points in one to have more clarity.	Pulse energy of 750 mJ or higher at 10 Hz should be demonstrated at the time of installation.
2	Beam Dimension ~ 24 x 10 mm ² Modified this criteria to accommodate maximum number of bidders.	Beam Dimension ~ height 24 to 28 mm and width 10 to 12 mm
3	Other necessary conditions: 1. Should have supplied minimum 5 similar high power Lasers in last five years in India. Should provide the list with addresses and contact details for these users. Modified this criteria to accommodate maximum number of bidders.	1. Should have supplied minimum 2-3 similar high power Lasers in India and minimum 10 worldwide during last five years. Should provide the list with addresses and contact details for these users.