

भारतीय विज्ञान शिक्षा एवं अनुसंधान संस्थान तिरुपती
INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH TIRUPATI

Transit campus: Sree Rama Engineering College Campus, Rami Reddy Nagar, Karakambadi Road, Mangalam
(B.O), Tirupati - 517 507



MENTORED BY

Indian Institute of Science Education and Research Pune
Dr. Homi Bhabha Raod, Pune : 411 008 (Maharashtra)
www.iiserpune.ac.in

INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH TIRUPATI

CLARIFICATION ON TENDER NUMBER - IISERT-PUR-0122-15

ITEM DESCRIPTION- PROCUREMENT OF SYRINGE PUMP

Refer our Press Tender Notice No.IISER/S&P/08/15-16 dated 24.11.2015 for procurement of **Syringe Pump**. Tender Reference Number - IISER-PUR-0122-15.

Pre-Bid meeting was held on December 04th , 2015 at 2.00 PM and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additons / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website www.iiserpune.ac.in will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

4.12.2015

Sd/-
Assistant Registrar (S&P)



IISER PUNE

PRE-BID CONFERENCE FOR PROCUREMENT OF SYRINGE PUMP

TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISERT-PUR-0122-15

DATE : 4.12.15

S.No	Query/Clarification Sought	Clarification / Amendment
1	Maximum Linear Force- 16 Kg (35lbs) @ 100 % force selection	At least 100 lbs @ 100% Force Selection
2	Type of sample	Water
3	Outline Of Application	Microfluidic research, which include pumping of fluids-mostly water based- through microchannels.



IISER PUNE

PRE-BID CONFERENCE FOR PROCUREMENT OF SYRINGE PUMP

COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISERT-PUR-0122-15

DATE : 4.12.15

S.No	Query/Clarification Sought	Clarification / Amendment
	-----NIL-----	-----NIL-----