



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH

PUNE

CLARIFICATION ON TENDER NUMBER - IISER-PUR-0321-14

ITEM DESCRIPTION- PROCUREMENT OF MULTIMODE MICROPLATE READER WITH ANALYSIS SOFTWARE

Refer our Press Tender Notice No.IISER/S&P/05/14 dated 8.7.2014 for procurement of Multimode Microplate Reader With Analysis Software.Tender Reference Number - IISER-PUR-0321-14.

Pre-Bid meeting was held on July 16th , 2014 at 10.30 AM and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website [www.iiserpune.ac .in](http://www.iiserpune.ac.in) will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

16.7.2014

Sd/-
Assistant Registrar (S&P)



IISER PUNE

PRE-BID CONFERENCE FOR PROCUREMENT MULTIMODE MICROPLATE READER WITH ANALYSIS SOFTWARE

TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-0321-14

DATE : 16.7.14

S.No	Query/Clarification Sought	Clarification / Amendment
1.	Whether separate cuvette port option is a must?	Yes, provision for cuvette based reading is absolutely must because the same instrument can be used for measuring absorbance or fluorescence of chemical compounds in organic solvents. It gives greater flexibility in handling chemistry and biology work
2.	Whether fluorescence polarization capability is a must?	Fluorescence polarization assay is one of the experimental techniques that my lab would use routinely? Therefore, it is very important that microplate reader has that capability
3.	Regarding capability to accommodate 12 × 75 mm test tubes?	System should have this feature, because it gives extra flexibility for the end user
4.	Is monochromator based detection system is a must or filter based detection okay?	Monochromator based detection is very essential because most of the times, we will be making new fluorphores where we may not know absorbance max and fluorescence max. Spectral scanning mode available

		in monochromator based system will be used to find Abs max and fluorescence max. we can't do this with filter based system
5.	Option for selection of band width is a must?	For a weakly fluorescent compounds, we might have to increase the bandwidth of emission filter in order to get good signal, therefore system with variable band width is preferred



IISER PUNE

PRE-BID CONFERENCE FOR PROCUREMENT MULTIMODE MICROPLATE READER WITH ANALYSIS SOFTWARE**COMMERCIAL QUERIES AND CLARIFICATION**

TENDER NUMBER - IISER-PUR-0321-14

DATE : 16.7.14

S.No	Query/Clarification Sought	Clarification / Amendment
	-----NIL-----	-----NIL-----