

भारतीय विज्ञान शिक्षा एवं अनुसंधान संस्थान तिरुपती
INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH TIRUPATI

Transit campus: Sree Rama Engineering College Campus, Rami Reddy Nagar, Karakambadi Road, Mangalam
(B.O), Tirupati - 517 507



MENTORED BY
Indian Institute of Science Education and Research Pune
Dr. Homi Bhabha Raod, Pune : 411 008 (Maharashtra)
www.iiserpune.ac.in

INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH TIRUPATI

CLARIFICATION ON TENDER NUMBER - IISERT-PUR-0109-15

ITEM DESCRIPTION- PROCUREMENT OF FREEZER DRYER

Refer our Press Tender Notice No.IISER/S&P/08/15 dated 24.11.2015 for procurement of Freezer Dryer. Tender Reference Number - IISER-PUR-0109-15.

Pre-Bid meeting was held on December 04th , 2015 at 2.00 PM and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additons / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website www.iiserpune.ac.in will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

4.12.2015

Sd/-
Assistant Registrar (S&P)



IISER PUNE

PRE-BID CONFERENCE FOR PROCUREMENT OF FREEZE DRYER

TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISERT-PUR-0109-15

DATE : 4.12.15

| S.No | Query/Clarification Sought | Clarification / Amendment |
|------|--|--|
| 1 | <p>Page No - 25, Chapter 4, Point2 Ice removal capacity in 24 hours: 3 Litres or better.</p> <p>Suggestion by one of the bidder "Kindly change the Ice removal capacity in 24 hours to 2.5 Liter or better in place of 3 liters or better"</p> | <p>Larger volume of water removal (3 Litres or more) is important in the context of smooth functioning of Freeze Dryer. Lowering this range to 2.5 litres in 24 hours is not possible.</p> <p>Hence, tender Specification prevails.</p> <p>No change in the specification.</p> |

| | | |
|---|---|--|
| 2 | <p>Page No - 25, Chapter 4, Point 3</p> <p>Overall Condenser Capacity of the system: 6-9 Litres.</p> <p>Data Acquisition: Up to 80 Hz</p> <p>Suggestion by one of the bidder: Overall condenser capacity 4.5 Liter in place of 6 to 9 Liters.</p> | <p>Larger capacity allows efficient freeze drying for multiple samples. Lowering this range to 4.5 litres is not possible.</p> <p>Hence, tender Specification prevails.</p> <p>No change in the specification.</p> |
| 3 | <p>Suggestion by one of the bidder: "Please mke provision of Hot gas Defrost as preferred feature"</p> | <p>This feature is not an essential feature required for our operations. However, if this provision is there in the instrument, the provision will comply.</p> |



IISER PUNE

PRE-BID CONFERENCE FOR PROCUREMENT OF FREEZE DRYER**COMMERCIAL QUERIES AND CLARIFICATION**

TENDER NUMBER - IISERT-PUR-0109-15

DATE : 4.12.15

| S.No | Query/Clarification Sought | Clarification / Amendment |
|------|----------------------------|---------------------------|
| | | |
| | -----NIL----- | -----NIL----- |
| | | |
| | | |
| | | |