



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH
PUNE

CLARIFICATION ON TENDER NUMBER - IISER-PUR-0010-17

ITEM DESCRIPTION- PROCUREMENT OF VIBRATION LESS CRYO-FREE
CRYOSTATION WITH ELECTROMAGNET

Refer our Press Tender Notice No.IISER/S&P/03/2017 dated 20.4.2017 for procurement of Vibration Less Cryo-free cryostation with Electromagnet Automated Protein Purification System. Tender Reference Number - IISER-PUR-0010-17.

Pre-Bid meeting was held on April 27th , 2017 at 3.30 PM and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website [www.iiserpune.ac .in](http://www.iiserpune.ac.in) will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

27.4.2017

Sd/-
Assistant Registrar (S&P)



IISER PUNE

**PRE-BID CONFERENCE FOR PROCUREMENT OF VIBRATION LESS CRYO-FREE
CRYOSTATION WITH ELECTROMAGNET**

TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-0010-17

DATE : 27.4.17

S.No	Query/Clarification Sought	Clarification / Amendment
1	Page No - 18, Chapter 4, Clause - 9 Cooldown time: “Cooldown time with the magnet will be unknown to us.”	Tender Specification prevails. No change in the specification.
2	Page No -18, Chapter 4, Clause - 10 Sample mount: “Are different sample holders being considered here for different	Page No -18, Chapter 4, Clause - 10 Sample mount:

	<p>distances from the windows?</p> <p>Mounting the sample holder at different angle might help to change the angle.</p> <p>Is manual change of distance and angle being considered here?</p> <p>If so, couple of different sample holders can be included.”</p>	<p>Manual change of distance and angle is considered here.</p>
3	<p>Page No - 18 , Chapter 4, Clause -11</p> <p>Positional flexibility:</p> <p>Do you plan to move vibration isolated cryostat too often? Please clarify.</p>	<p>Page No - 18 , Chapter 4, Clause -11</p> <p>Positional flexibility:</p> <p>Yes. We would like to use the cryostat for three different experimental set ups built on the same optical table.</p>
4	<p>Page No - 18, Chapter - 4, Clause -15</p> <p>Software and laptop:</p> <p>Generally we do not provide laptop and software. Generally the researchers make their own software depending on their need. What software are you looking for here?</p>	<p>Page No - 18, Chapter - 4, Clause -15</p> <p>Software and laptop:</p> <p>The software is to control the operation of the cryostat and the magnet. The vendor has to provide the laptop. In addition, the vendor should also provide the related Labview drivers and sample Labview codes.</p>

5	<p>Page No - 18, Chapter 4, Clause - 16</p> <p>Compressor:</p> <p>We do not offer Variable flow compressor.</p> <p>We can offer air cooled compressor from Sumitomo, but this will not have variable flow.</p>	<p>Page No - 18, Chapter 4, Clause - 16</p> <p>Compressor:</p> <p>Tender Specification prevails.</p> <p>No change in the specification.</p>
6	<p>Page No - 18 , Chapter 4, Clause - 18</p> <p>Remote trouble shoot:</p> <p>We do not offer an option with capability of trouble shooting from remote.</p>	<p>Page No - 18 , Chapter 4, Clause - 18</p> <p>Remote trouble shoot:</p> <p>Tender Specification prevails.</p> <p>No change in the specification.</p>
7	<p>Page No - 19 , Chapter 4</p> <p>Electromagnet:</p> <p>We prefer the customer to get magnet separately, not from us. If we really have to source a magnet, we need to get a quote to meet the customer's specifications.</p>	<p>Page No - 19 , Chapter 4</p> <p>Electromagnet:</p> <p>The vendor should provide electromagnet (1 Tesla) with the cryostat.</p>



IISER PUNE

**PRE-BID CONFERENCE FOR PROCUREMENT OF VIBRATION LESS CRYO-FREE
CRYOSTATION WITH ELECTROMAGNET**

COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-0010-17

DATE : 27.4.17

S.No	Query/Clarification Sought	Clarification / Amendment
	-----NIL-----	-----NIL-----