



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH

PUNE

CLARIFICATION ON TENDER NUMBER - IISER-PUR-1279-13

ITEM DESCRIPTION- PROCUREMENT OF GLOVE BOX WITH 3 PORTS

Please refer our Press Tender Notice No.IISER/S&P/14/13 dated 10.01.2014 for procurement of Glove Box With 3 Ports . Tender Reference Number - IISER-PUR-1279-13.

Pre-Bid meeting was held on 24th January 2014 at 11.00 am and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website [www.iiserpune.ac .in](http://www.iiserpune.ac.in) will remain unchanged.No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair.

24.1.2014

Sd/-
Assistant Registrar (S&P)



IISER PUNE

PRE-BID CONFERENCE FOR PROCUREMENT OF GLOVE BOX WITH 3 PORTS

TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-1279-13

DATE : 24.1.14

S.No	Query/Clarification Sought	Clarification / Amendment
1	Chapter 4; Page 23 Query: IISER tender states Automatic pressure control with no manual intervention.	Chapter 4; Page 23 As per the Pre-bid discussion, this has been changed to Automatic pressure control along with manual pedal control.
2	Chapter 4; Page 24 Query: IISER tender states highly efficient vacuum pump with flow more than 20 Cu.m/hr. Can this be reduced?	Chapter 4; Page 24 The minimum flow rate should at least be between 12 and 15 Cu,m/hr
3	Chapter 4; Page 24 Query: The IISER tender did not specify the blower speed in the purification section. Can this be specified?	Chapter 4; Page 24 The maximum flow rate required is in the range of 95-110 m ³ /h (CFM). These flow rates should be variable as per various purification and handling needs.
4	Chapter 4, Page 24 Query: IISER tender states the sound level should be below 49 dB. Can this be higher?	Chapter 4, Page 24 The sound level can be at a maximum limit of 60 dB.
5	Chapter 4, Page 25 Query: IISER tender states the O ₂ and H ₂ O analyzers should be in the measuring range of 0-23000 ppm. Can the maximum detection limit be changed?	Chapter 4, Page 25 This range can be changed to 0-1000 ppm and higher.



IISER PUNE

PRE-BID CONFERENCE FOR PROCUREMENT OF GLOVE BOX WITH 3 PORTS

COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-1279-13

DATE : 24.1.14

S.No	Query/Clarification Sought	Clarification / Amendment
1.	<p>Security deposit/ Performance bank Guarantee:</p> <p>IISER Tender states Bank Guarantee is required for 10%. Can this replaced with 90% Sight draft and balance 10% through TT after completion of Warranty.</p>	<p>Security deposit/ Performance bank Guarantee</p> <p>No change in tender terms & conditions</p>