



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH
PUNE

CLARIFICATION ON TENDER NUMBER - IISER-PUR-1678-14

ITEM DESCRIPTION- PROCUREMENT OF RHEOMETER FOR POLYMER ANALYSIS

Refer our Press Tender Notice No.IISER/S&P/17/14 dated 2.3.2015 for procurement of Rheometer For Polymer Analysis . Tender Reference Number - IISER-PUR-1678-14.

Pre-Bid meeting was held on March 11th, 2015 at 2.00 PM and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website www.iiserpune.ac.in will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

11.3.2015

Sd/-
Assistant Registrar (S&P)



IISER PUNE

PRE-BID CONFERENCE FOR PROCUREMENT OF RHEOMETER FOR POLYMER ANALYSIS

TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-1678-14

DATE : 11.3.15

S.No	Query/Clarification Sought	Clarification / Amendment
1	Tender specification, point 3: Torque Range: 0.5 nNm to 200 mNm Change Requested for: Torque Range: 3.0 nNm to 200 mNm	No Change in Tender specifications
2	Tender specification, point 4: Torque Oscillation: minimum 2 nNm Change Requested for: 3 nNm	No Change in Tender specifications
3	Tender specification, point 16: Small Angle Light Scattering accessories Change Requested for: Microscopy and FT-IR accessories	No Change in Tender specifications

4	<p>Tender specification, point 8: <u>Frequency Range</u> Minimum: 1.0 μHz Maximum: 150 Hz</p> <p><u>Change Requested for:</u> Frequency Range Minimum: 1.0 μHz Maximum: 100 Hz</p>	<p>Tender Specification Amended as</p> <p><u>Frequency Range changed into</u></p> <p>Minimum: 1.0 μHz Maximum: 100 Hz</p>



IISER PUNE

PRE-BID CONFERENCE FOR PROCUREMENT OF RHEOMETER FOR POLYMER ANALYSIS
COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-1678-14

DATE : 11.3.15

S.No	Query/Clarification Sought	Clarification / Amendment
	-----NIL-----	-----NIL-----