



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH PUNE

CLARIFICATION ON TENDER NUMBER - IISER-PUR-0629-16

ITEM DESCRIPTION- PROCUREMENT OF **Plant Growth Chamber - 02 Units**

Refer our Press Tender Notice No.IISER/S&P/11/2016 dated 03.10.2016 for procurement of **Plant Growth Chamber - 02 Units**. Tender Reference Number - IISER-PUR-0629-16.

Pre-Bid meeting was held on October 14th, 2016 at 02.00 PM and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website [www.iiserpune.ac .in](http://www.iiserpune.ac.in) will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

17.10.2016

Sd/-
Assistant Registrar (S&P)



IISER PUNE

PRE-BID CONFERENCE FOR PROCUREMENT OF PLANT GROWTH CHAMBER--02

TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-0629-16

DATE : 00.00.16

S.No	Query/Clarification Sought	Clarification / Amendment
1	Page No - 24, "Common requirements or specifications for Unit 1 and 2" Point 3: Are the plant growth chambers with 10% higher in external dimensions acceptable?	Tender Specification prevails. No change in the specification.
2	Page No - 24, "Common requirements or specifications for Unit 1 and 2": Are the plant growth chambers with 10% lower in internal dimensions (the actual plant growth space) acceptable?	Tender Specification prevails. No change in the specification.
3	Page No - 24, "Common requirements or specifications for Unit 1 and 2" Point 9: Is spray humidifier in place of an ultrasonic one acceptable?	Tender Specification prevails. No change in the specification.
4	Page No - 24, "Common requirements or specifications for Unit 1 and 2" Point 10: Is fluorescent lighting in place of LED acceptable?	Tender Specification prevails. No change in the specification.
5	Page No - 24, "Common requirements or specifications for Unit 1 and 2" Point 11: Is programming memory of 4-5 programs acceptable?	Tender Specification prevails. No change in the specification.
6	Page No - 24, "Common requirements or specifications for Unit 1 and 2"	Tender specification is amended according to this

	Point 9: Please consider “irrespective of the ambient temperature fluctuations” in place of “irrespective of outside weather condition”.	suggestion; please read “irrespective of outside weather condition” as “irrespective of the ambient temperature fluctuations”.
7	Page No - 24, “Common requirements or specifications for Unit 1 and 2” Point 10: Are LEDs other than white-light ones acceptable if their light’s intensity is lower than 700 $\mu\text{mol}/\text{m}^2/\text{s}$?	Tender Specification prevails. No change in the specification.
8	Page No - 25, “Common requirements or specifications for Unit 1 and 2” Point 17: Since recently the entire institute’s power supply has been stabilized, is an additional voltage stabilizer necessary?	Stabilizer can be quoted separately.
9	Whether warranty should be quoted separately or included with the plant growth chamber cost?	Totals of plant growth chamber and warranty quotes from different bidders will be compared. So quotes with included warranty will be convenient for such comparison.
10.	Page No - 24, “Unit 2: Plant growth chamber having two compartments (>650L volume each) with independent light, temperature and humidity controls.” Are two different small units in place of one with two independent compartments acceptable?	One chamber with two compartments is preferable; however, two different small units will also be considered, provided they do not occupy extra space when stacked or kept side-by-side.



ANNEXURE -III

IISER PUNE

PRE-BID CONFERENCE FOR PROCUREMENT OF PLANT GROWTH CHAMBER--02

COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-0629-16

DATE : 14.10.16

S.No	Query/Clarification Sought	Clarification / Amendment
	-----NIL-----	-----NIL-----