



## INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH

PUNE

CLARIFICATION ON TENDER NUMBER - IISER-PUR-0148-15

ITEM DESCRIPTION- PROCUREMENT OF GLOVE BOX

Refer our Press Tender Notice No-IISER/S&P/04/15-16 dated 25.6.2015 for procurement of Glove Box. Tender Reference Number - IISER-PUR-0148-15.

Pre-Bid meeting was held on July 03<sup>rd</sup> , 2015 at 2.00 PM and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website [www.iiserpune.ac .in](http://www.iiserpune.ac.in) will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

3.7.2015

Sd/-  
Assistant Registrar (S&P)



## IISER PUNE

**PRE-BID CONFERENCE FOR PROCUREMENT OF GLOVE BOX**

## TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-0148-15

DATE : 3.7.15

S.No	Query/Clarification Sought	Clarification / Amendment
1	IISER tender states Glove ring diameter should be 220 mm. Although the standard catalogue of the product does not have this specification, would it be acceptable if this can be quoted in the technical bid as per the requirement?	We will accept the customized specification of 220 mm diameter from the vendors. However, the dealer(s) must provide a certification from the principals that they will be able to customize and provide the required specification for Glove rings.
2	IISER tender states the Box-Light should be LED lamp and will be inside the box. Although the standard catalogue of the product mentions about the fluorescent lamp, would it be acceptable if LED lamp can be quoted in the technical bid as per the requirement? Also, would it be acceptable if the Box-lamp is located outside the box?	We will accept the customized specification of LED-lamp from the vendors. However, the dealer(s) must provide a certification from the principals that they will be able to customize and provide the required specification for the box-lamp. The box-lamps can be located outside the box, but they should be protected by a separate glass-cover.

3	The IISER tender specifies the mini-antechamber dimensions of 150(D) x 400(L) mm. Although the standard catalogue of the product mentions a length of 300 mm, would it be acceptable if the 400 mm length of the mini-antechamber can be quoted in the technical bid as a custom specification?	We will accept the customized specification of 400 mm length for the mini-antechamber. However, the dealer(s) must provide a certification from the principals that they will be able to customize and provide the required specification for the mini-antechamber dimensions.
4	IISER tender did not mention the full balance specifications and model. Would the price of the balance affect the commercial evaluation of the bids?	The balance is an accessory item and its price will not affect the commercial comparisons. However, we request the vendors to quote at least 2-3 models with separate technical details for each of them.
5	IISER tender states the O <sub>2</sub> analyzer should be in the measuring range of 0->1000 ppm. Can the maximum detection limit be changed?	This range can be changed to 0-1000 ppm and higher.



IISER PUNE

**PRE-BID CONFERENCE FOR PROCUREMENT OF GLOVE BOX**  
COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-0148-15

DATE : 3.7.15

S.No	Query/Clarification Sought	Clarification / Amendment
1.	Whether the Name of Equipment is Glow Box or Glove Box	Please read the name of Equipment as <b>Glove Box</b>