



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH

PUNE

CLARIFICATION ON TENDER NUMBER - IISER-PUR-0204-14

ITEM DESCRIPTION- PROCUREMENT OF AUTOMATED PROTEIN PURIFICATION SYSTEM

Refer our Press Tender Notice No.IISER/S&P/05/14 dated 8.7.2014 for procurement of Automated Protein Purification System. Tender Reference Number - IISER-PUR-0204-14.

Pre-Bid meeting was held on July 16th , 2014 at 10.30 AM and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website [www.iiserpune.ac .in](http://www.iiserpune.ac.in) will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

16.7.2014

Sd/-
Assistant Registrar (S&P)



IISER PUNE

PRE-BID CONFERENCE FOR PROCUREMENT OF AUTOMATED PROTEIN PURIFICATION SYSTEM

TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-0204-14

DATE : 16.7.14

S.No	Query/Clarification Sought	Clarification / Amendment
1	Is flow rate of 10 ml/min okay?	Flow rate up to 25 ml/min is very essential for purification of large quantity of protein in a single run. We can't do this if system has capability only go to 10 ml/min
2	Automatic gradient formation from 0 to 100% gradient over the entire flow range of 0.1 to 25 ml/min. is this feature must?	The system should definitely have this feature, otherwise it will be difficult for us to use the machine when we do ion exchange chromatography or hydrophobic interaction chromatography
3.	Is a cold cabinet should be provided by the same vendor who supply the instrument?	This is not an absolute must however cold cabinets provided by the third party vendors may not best suit the instrument accommodation. This is why I prefer to buy the cold cabinet from the same vendor



IISER PUNE

**PRE-BID CONFERENCE FOR PROCUREMENT OF AUTOMATED PROTEIN PURIFICATION
SYSTEM**

COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-0204-14

DATE : 16.7.14

S.No	Query/Clarification Sought	Clarification / Amendment
	-----NIL-----	-----NIL-----