



## INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH

PUNE

CLARIFICATION ON TENDER NUMBER - IISER-PUR-0650-13

ITEM DESCRIPTION- PROCUREMENT OF GLOW BOX WITH SOLVENT PURIFICATION

Please refer our Press Tender Notice No.IISER/S&P/06/13 dated 09<sup>th</sup> August 2013 for procurement of Glow Box With Solvent Purification. Tender Reference Number - IISER-PUR-0650-13.

Pre-Bid meeting was held on 20<sup>th</sup> August 2013 at 2.00 PM and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied there to as detailed in Annexure -II

The following points are noted.

1. No written queries were received from any bidder.
2. Under Chapter - 5, page no - 24, Under Heading Gas Purification Unit , the Sub Heading Leakage rate is deleted and added in page no 23, Under Heading Antechamber.
3. In general the bidder expressed their satisfaction about the details and specifications given in the bid document.

It was informed that the corrections / additons / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website [www.iiserpune.ac .in](http://www.iiserpune.ac.in) will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair.

20.8.2013

Sd/-  
Assistant Registrar (S&P)



## IISER PUNE

### PRE-BID CONFERENCE FOR PROCUREMENT OF HIGH PERFORMANCE STEADY STATE FLUORIMETER

#### TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-0650-13

DATE : 20.08.13

S.No	Query/Clarification Sought	Clarification / Amendment
1.	<p>Page no - 23, Chapter- 4, Under Schedule of Requirement, Specifications &amp; Allied Technical Details</p> <p>Please refer Page no -24 , Under Heading <b><u>Gas Purification Unit</u></b></p> <p><b>Leakage Rate:</b> &lt;10-5 mbar l/s (ISO complaint)</p>	<p>Page no - 23, Chapter- 4, Under Schedule of Requirement, Specifications &amp; Allied Technical Details</p> <p><b>Leakage Rate : : &lt;10-5 mbar l/s (ISO complaint)</b></p> <ul style="list-style-type: none"><li>• <b>Added</b> in Page no - 23, Under Heading <b><u>Antechamber</u></b></li><li>• <b>Deleted from</b> Page no -24 , Under Heading <b><u>Gas Purification Unit</u></b></li></ul>



## IISER PUNE

## PRE-BID CONFERENCE FOR PROCUREMENT OF HIGH PERFORMANCE STEADY STATE FLUORIMETER

## COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-0650-13

DATE :20.8.13

S.No	Query/Clarification Sought	Clarification / Amendment
	----- NIL -----	-----NIL-----