



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH PUNE

CLARIFICATION ON TENDER NUMBER - IISER-PUR-1544-20

ITEM DESCRIPTION- PROCUREMENT OF DUAL BEAM SYSTEM COMPRISING OF FIB-FE SEM MACHINE

Refer IISER Pune global tender notice published in Indian Express Pune, Mumbai and New Delhi editions for procurement of Dual beam system comprising of FIB-FE SEM

Pre-Bid meeting was held on June 29th, 2021 at 3.00 PM via video conferencing and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website [www.iiserpune.ac .in](http://www.iiserpune.ac.in) will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

29.6.2021

Sd/-
Assistant Registrar (S&P)



IISER PUNE

PRE-BID CONFERENCE FOR PROCUREMENT OF DUAL BEAM SYSTEM COMPRISING OF FIB-FE SEM

TECHNICAL & COMMERCIAL QUERIES AND CLARIFICATION

S.No	Query/Clarification Sought	Clarification / Amendment
1	We request at least 1.0 nm at 15 kV and 1.4 nm at 1 kV.	Resolution At least 1nm at 15kV and 1.8nm at 1kV. Higher kV resolution has been changed as suggested which is very close to our requirement. Lower kV resolution is kept unchanged as that is sufficient for our purpose.
2	We request an amendment in spec to either a electrostatic or electromagnetic or combo of both.	No change in tender specifications. This is required for our experiment.
3	We request an amendment in beam current beam current 1pA-65 nA	No change in tender specifications. This is required for our experiment.
4	Stability of beam 0.4 % for 10 hours of measurement. Added we would like to clarify that; we don't have this specification in our datasheet. Will a confirmation on letter head suffice to requirements.	No change in tender specifications. Yes, the stability has to be shown on-site.