



## **INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH PUNE**

CLARIFICATION ON TENDER NUMBER - IISER-PUR-0743-18

### **ITEM DESCRIPTION- PROCUREMENT OF MICROSCOPE AND CONFOCAL SCAN HEAD**

Please refer our Press Tender Notice No.IISER/S&P/09/18 dated 27.10.2018 for procurement of **MICROSCOPE AND CONFOCAL SCAN HEAD**. In this regard refer following.

1. Pre-Bid Conference dated November 02<sup>nd</sup>, 2018
2. Corrigendum on Specifications - As per Annexure A

**Please note that there is no change in last date and time for submission of bid documents. The last date remains same i.e November 20<sup>th</sup>, 2018, Time - 3.00 pm.**

All the Prospective Bidders are required to take cognizance of corrigendum along with proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding

The other terms & conditions of the notice issued on our IISER website [www.iiserpune.ac.in](http://www.iiserpune.ac.in) will remain unchanged.No more correspondence in this regard will be entertained

16.11.2018

Sd/-  
Assistant Registrar (S&P)

## Annexure A

### Changes made in specifications

#### Microscope specifications

Item No. 2: Motorised XY stage for XY scanning fitted with XZ Piezo/Galvo stage with a travel range of 100 micron or better. (Galvo has been added).

#### Spectral confocal imaging unit with laser module

Item No 1 : It should be compatible with Leica DMI-8. This sentence has been deleted.

Item No. 1 : Laser Unit including following laser lines with laser power :405/408nm, 635nm/640nm (optional) with a power of 20 mW or better. (Optional is added)

#### Other requirements

Item No 3 : Computer table and laser rack. (This item has been added.)