



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH
PUNE

CLARIFICATION ON TENDER NUMBER - IISER-PUR-171-20

ITEM DESCRIPTION- PROCUREMENT OF ISTOTHERMAL CALORIMETER

Refer IISER Pune open tender number IISER-PUR-123-20 dated 9.7.2020 for procurement of Isothermal Calorimeter.

Pre-Bid meeting was held on July 30th , 2020 at 3.00 PM via video conferencing and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additons / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website [www.iiserpune.ac .in](http://www.iiserpune.ac.in) will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

30.7.2020

Sd/-
Assistant Registrar (S&P)



IISER PUNE

PRE-BID CONFERENCE FOR PROCUREMENT OF ISOTHERMAL CALORIMETER

TECHNICAL & COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-0171-20

DATE : 30.7.2020

S.No	Query/Clarification Sought	Clarification / Amendment
1	Earnest Money Deposit Page 1 of the Tender Document / Tender Notice lists the Earnest Money Deposit to Be Rs 150,000 and Page 6 (Chapter 1) lists the amount to be Rs. 175,000. We request confirmation on what is the correct amount to be considered for submission	Earnest Money Deposit The Earnest Money Deposit is to be deposited for Rs 1,50,000/- .
2	Warranty The tender requires a warranty of 5 years. We request that the warranty be limited to a period of 3 years instead of 5 years. Post 3 year warranty period Annual Maintenance Services can be offered as per the tender documents. We have close to 150 installations of products in India and they have proven reliability over many years of trouble free operation.	Warranty clause amended to : The warranty period is amended to 3 years period.

3	<p>Security Deposit / Performance Bank Guarantee Performance Bank Guarantee (PBG) will be issued for 10% of the order value valid for 36 months + 60 days only. The said PBG will be issued by principal through their above-mentioned bank. Our principal will not issue a PBG for a longer period and we request your acceptance for this and submission of PBG only after completion of installation.</p>	<p>Security Deposit / Performance Bank Guarantee</p> <p>The PBG is to be issued for 3 years period routed through Indian Bank.</p>
4	<p>“Our instrument model have different specifications and we are sure these would support your end use applications in a better way. We would like to highlight the deviations as given below.</p> <p>1) S. No. 3 Injection parameters a) Injection syringe capacity: The injection syringe capacity should be $\leq 50 \mu\text{L}$ or less. The system should have bubble free sample loading that can take care of high viscous solutions.</p> <p>2) S. No. 3 Injection parameters b) Mixing or stirring speed: The system should have user definable mixing speed options (rotations per minutes, rpm) with maximum stirring speed up above 1500 rpm.</p> <p>S. No. 4 a) Response time: The system should be capable to provide user selectable response times with a minimum response time $\sim \leq 8$ seconds.</p> <p>S. No b) User selectable Mode of operation for response times: Multiple mode option for response times (high, medium and low gains) should be available to cover broader range of binding reactions, and allow fast re-equilibration between injections.”</p>	<p>No Change in tender specification</p> <p>Tender specification prevails</p>
5	<p>Page No - 24, Chapter 4</p> <p>Necessary Accessories: Accessories for sample degassing should be provided or options based on the system configurations for avoiding air bubbles during operation should be specified.</p>	<p>Page No - 24, Chapter 4</p> <p>Necessary Accessories - The sample degassing should be quoted separately as optional item.</p>

	<p>TA Instruments Submission: It is not clear if the system needs to be quoted along with the sample degassing accessory. In every ITC experiment the sample needs to be degassed before the experiment. It is necessary that everyone quote their degassing accessory.</p>	<p>While evaluating L1 the optional accessories will not be considered.</p>
--	---	---