

FILE NO - IISER-PUR-0272-13

Minutes of the meeting of the Pre-Bid Conference for the Procurement of High Performance Steady State Fluorimeter held on 24th June 2013 at 4.00 PM in the Board Room of IISER

The Pre-Bid Conference was held and the following members attended the meeting:

1. Dr. Girish Ratnaparkhi - Chairman
2. Dr. Thomas Pucadyil
3. Dr. Partha Hazra
4. Dr. Saikrishnan Kayarat
5. Mr. Salim Shaikh

The list of prospective bidders who attended the meeting as per Annex -1

A press tender notice was issued on 15th June 2013 for the procurement of High Performance Steady State Fluorimeter under two-bid system.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied there to as detailed in Annexure -II


The representatives present were satisfied with the replies given and it was informed that the corrections / additons / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website www.iiserpune.ac.in will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair.



Dr. Girish Ratnaparkhi



Dr. Thomas Pucadyil



Dr. Partha Hazra



Dr. Saikrishnan Kayarat



Mr. Salim Shaikh



IISER PUNE

PRE-BID CONFERENCE FOR PROCUREMENT OF HIGH PERFORMANCE STEADY STATE FLUORIMETER

ATTENDENCE SHEET OF PROSPECTIVE BIDDERS

TENDER NUMBER - IISER-PUR-0272-13

DATE : 24.06.13

S.No	NAME OF THE FIRM	REPRESENTED BY	LANDLINE & MOBILE NO	E-MAIL ID	SIGNATURE	REMARKS
1	Advantech Ind.	Dr. A. K. Maiti	9004875731	smaiti65@gmail.com	A. Maiti	

Girish Ratnaparkhi
Dr. Girish Ratnaparkhi

Thomas Pucadyil
Dr. Thomas Pucadyil

Partha Hazra
Dr. Partha Hazra

DR. PARTHA HAZRA

Saikrishnan Kayarat
Dr. Saikrishnan Kayarat

Salim Shaikh
Mr. Salim Shaikh



IISER PUNE

PRE-BID CONFERENCE FOR PROCUREMENT OF HIGH PERFORMANCE STEADY STATE FLUORIMETER

TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-0272-13

DATE : 24.06.13

S.No	Query/Clarification Sought	Clarification / Amendment
1.	Page 23, Section LIGHT SOURCE. Is vertically mounted lamp a requirement?	Vertically mounted lamp is not necessary, other formats are permissible
2.	Page 23, Section SAMPLE COMPARTMENT. Should the sample compartment be equipped with a temperature controller unit	Sample compartment should be equipped with a dedicated temperature controlled unit.

Dr. Girish Ratnaparkhi

Dr. Thomas Pucadyil

Dr. Partha Hazra

Dr. Saikrishnan Kayarat

Mr. Salim Shaikh



IISER PUNE

PRE-BID CONFERENCE FOR PROCUREMENT OF HIGH PERFORMANCE STEADY STATE FLUORIMETER

COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-0272-13

DATE : 24.06.13

S.No	Query/Clarification Sought	Clarification / Amendment
	— NIL —	

Dr. Girish Ratnaparkhi

Dr. Thomas Pucadyil

Dr. Partha Hazra

Dr. Saikrishnan Kayarat

Mr. Salim Shaikh