



Bid Number: GEM/2020/B/671427
Dated: 12-06-2020

Bid Document

Bid Details	
Bid End Date/Time	25-06-2020 11:00:00
Bid Opening Date/Time	2020-06-25 11:30:00
Bid Life Cycle (From Publish Date)	90 (Days)
Bid Offer Validity (From End Date)	30 (Days)
Ministry/State Name	Ministry Of Human Resource Development
Department Name	Department Of Higher Education
Organisation Name	Indian Institute Of Science Education And Research (iisers)
Office Name	Pune 411008
Total Quantity	20
Item Category	digital multimeter
MSE Exemption for Years of Experience and Turnover	No
Startup Exemption for Years of Experience and Turnover	No
Document required from seller	Certificate (Requested in ATC),OEM Authorization Certificate *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Bid to RA enabled	No
Estimated Bid Value	200000

EMD Detail

Required	No
----------	----

ePBG Detail

Required	No
----------	----

Splitting

Bid splitting not applied.

Digital Multimeter (20 pieces)

Technical Specifications

[* As per GeM Category Specification](#)

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
FEATURES	Type	Hand Held	Hand Held
	Display (Digits)	3.5	3.5
	Measurement Type	Average	Average
	Counts	4000	4000, 6000, 50000, 10000, 20000, 30000, 40000, 60000, 70000, 80000, 90000, 100000, 200000 Or higher
MEASUREMENTS	DC Voltage Range	1000	1000.0
	DC Voltage Accuracy	0.1	0.1
	DC Voltage resolution	100	100.0
	AC Voltage Range	1000	1000.0
	AC Voltage Accuracy	1	1.0
	AC Voltage resolution	100	100.0
	DC Current Range	10	10.0
	DC Current Accuracy	1.5	1.5
	DC Current resolution	1	1.0
	AC Current Range	10	10.0
	AC Current Accuracy	1.5	1.5
	AC Current resolution	1	1.0
	Resistance Range	100	100.0
	Resistance Accuracy	3.5	3.5
	Resistance resolution	10	10.0
	Capacitance Range (Micro farad)	1 to 200	1 to 200
	Capacitance Accuracy (%)	3.5	3.5
Capacitance resolution (nF)	10	10	
Frequency Range (200	200	

	KHz)		
	Frequency Accuracy (%)	1	1
	Frequency resolution (Hz)	10	10
	Temperature Range (Degree C)	-40 to 400	-40 to 400
	Temperature Accuracy (%)	1.5	1.5
	Temperature resolution (Degree C)	0.001	0.001
OTHER FEATURES	Continuity test and Diode test	Available	*
	Data logging	Available	*
	If Data Logging Feature is Available, Internal Memory	min. 5000 points/ 200 readings	*
	Computer Interface along with software	Not Available	*
	Dimensions (mm X mm X mm)	-	*
	Weight	88	*
	Other Features	-	*
SAFETY	Safety Rating	Category II 600 V of IEC 61010.	Category II 600 V of IEC 61010.
	Safety Certification	UL,CE,CSA,TUV,VDE	*
POWER REQUIREMENTS	Power Requirements	Battery Operated	*
	if Battery Operated, then Type of Battery	Non- Rechargeable (Alkaline)	*
	if rechargeable , then chemistry of the battery	NA	*
	if Non-rechargeable , then size of the battery / cell	AAA	*
	number of non-rechargeable the battery / cell	2	*
	Voltage of rechargeable Battery	9	*
	capacity of rechargeable battery(milli Ampere Hour)	2500	*

	Suitable charger / adaptor provided	YES	*
ACCESSORIES	Standard Accessories Supplied with the Equipment	-	*
WARRANTY	Warranty	1	1.0
OPERATING CONDITION	Minimum Operating Temperature	-10	*
	Maximum Operating Condition	60	*
	Operating Humidity (RH)	85	*
TEST REPORT DETAILS	The equipment supplied with calibration certificate	YES	*
	conformity to Dry Heat for 16 hrs at a specified maximum temp in accordance with IS:9000/part-3/section-5/1977 latest	Yes	*
	conformity to Cold Test for 4 hrs at a specified minimum temp in accordance with IS:9000/part-2/section-4/1977 latest	Yes	*
	conformity to Damp Heat (Cyclic)Test for 2 Cycles of 24 hrs each at a temp of 40 degree C & 95% RH	Yes	*
	Availability of Test Reports from Central Govt/NABL/ILAC approved accredited lab to prove conformity to the specification	Yes	*
	Test Report Number and date for the specification test	-	*
	Name and address of the Lab for the specification test	-	*

* Specifications highlighted in bold are the Golden Parameters.

* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement (allowed Values) by the Buyer.

Additional Specification Documents

Consignees/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1	John Albert K	411008,Indian Institute of Science Education and Research, Pune (IISER) Dr.Homi Bhabha Road , Pashan Pune 411008 Email: purchase@iiserpune.ac.in For Payment Status: 020 2590 8180/8452 For Delivery Status : 020 2590 8246, PUNE CITY, 411008	20	30

Bid Specific Additional Terms and Conditions

1. OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity by up to 25% of the contracted quantity during the currency of the contract at the contracted rates. Bidders are bound to accept the orders accordingly.
2. The buyer organization is an institution eligible for concessional rates of GST as notified by the Government of India. The goods for which bids have been invited fall under classification of GST concession and the conditions for eligibility of concession are met by the institution. A certificate to this effect will be issued by Buyer to the Seller after award of the Contract. Sellers are requested to submit their bids after accounting for the Concessional rate of GST. Applicable Concessional rate of GST : 5%
Notification No.and date : The rate of GST application to IISER Pune is 5% for the items procured for Research purpose as per notification No 45/2017 - Central Tax (Rate) New Delhi. dated 14/11/2017
3. Timely Servicing / rectification of defects during warranty period: After having been notified of the defects / service requirement during warranty period, Seller has to complete the required Service / Rectification within 10 days time limit. If the Seller fails to complete service / rectification with defined time limit, a penalty of 0.5% of Unit Price of the product shall be charged as penalty for each week of delay from the seller. Seller can deposit the penalty with the Buyer directly else the Buyer shall have a right to recover all such penalty amount from the Performance Security (PBG).Cumulative Penalty cannot exceed more than 10% of the total contract value after which the Buyer shall have the right to get the service / rectification done from alternate sources at the risk and cost of the Seller besides forfeiture of PBG. Seller shall be liable to re-imburse the cost of such service / rectification to the Buyer.
4. Scope of supply (Bid price to include all cost components) : Supply Installation Testing and Commissioning of Goods
5. Dedicated /toll Free Telephone No. for Service Support : BIDDER/OEM must have Dedicated/toll Free Telephone No. for Service Support.

[This Bid is also governed by the General Terms and Conditions](#)

---Thank You---