

भारतीय विज्ञान शिक्षा एवं अनुसंधान संस्थान तिरुपती

INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH TIRUPATI

**Transit campus:** Sree Rama Engineering College Campus, Rami Reddy Nagar, Karakambadi Road, Mangalam (B.O), Tirupati - 517 507



MENTORED BY

Indian Institute of Science Education and Research Pune

Dr. Homi Bhabha Raod, Pune : 411 008 (Maharashtra)

[www.iiserpune.ac.in](http://www.iiserpune.ac.in)

INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH TIRUPATI

CLARIFICATION ON TENDER NUMBER - IISERT-PUR-0127-15

ITEM DESCRIPTION- PROCUREMENT OF AUTOMATIC DIGITAL POLARIMETER

Refer our Press Tender Notice No.IISER/S&P/08/15-16 dated 24.11.2015 for procurement of **Automatic Digital Polari meter** . Tender Reference Number - IISERT-PUR-0127-15.

Pre-Bid meeting was held on December 04<sup>th</sup> , 2015 at 2.00 PM and minutes of meeting is as under .

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additons / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website [www.iiserpune.ac.in](http://www.iiserpune.ac.in) will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

4.12.2015

Sd/-  
Assistant Registrar (S&P)



## IISER PUNE

**PRE-BID CONFERENCE FOR PROCUREMENT OF AUTOMATIC DIGITAL POLARIMETER**

## TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISERT-PUR-0127-15

DATE : 4.12.15

S.No	Query/Clarification Sought	Clarification / Amendment
1	Page No - 25, Chapter 4, Temperature probe range and accuracy Suggestion by one of the bidders "We request to include Temperature control by Peltier Thermostat"	The instrument should have Temp. Probe Range : 10°-100°C Temp. Probe Accuracy : ±0.1°C The control by Peltier Thermostat is compulsory.
2	Page No - 25, Chapter 4 Suggestion by one of the bidders: Sample cell path length 100 mm, sample cell material: Stainless steel (for faster temperature control)	Sample cell path length of 100 mm will comply. Stainless steel material is not essential feature for our operations.

3	<p>Page No - 25, Chapter 4</p> <p>Suggestion by one of the bidders: Include the point "A Suitable PC should be Supplied for data storage"</p>	<p>Suitable latest PC should be quoted under optional items.</p>
4	<p>Suggestion by one of the bidders: Include the point "A Provision to detect air bubble in the SS sample cell should be quoted"</p>	<p>This feature is not essential for our Teaching and Research.</p> <p>Tender specification prevails</p>



IISER PUNE

**PRE-BID CONFERENCE FOR PROCUREMENT OF AUTOMATIC DIGITAL POLARIMETER****COMMERCIAL QUERIES AND CLARIFICATION**

TENDER NUMBER - IISERT-PUR-0127-15

DATE : 4.12.15

S.No	Query/Clarification Sought	Clarification / Amendment
	-----NIL-----	-----NIL-----