



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH  
PUNE

CLARIFICATION ON TENDER NUMBER - IISER-PUR-1020-14

ITEM DESCRIPTION- PROCUREMENT OF CRYOGENIC PROBE STATION WITH ACCESSORIES

Refer our Press Tender Notice No.IISER/S&P/9/14 dated 24.11.2014 for procurement of Cryogenic Probe Station with Accessories. Tender Reference Number - IISER-PUR-1020-14.

Pre-Bid meeting was held on December 03<sup>rd</sup> , 2014 at 10.30 AM and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website [www.iiserpune.ac .in](http://www.iiserpune.ac.in) will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

3.12.2014

Sd/-  
Assistant Registrar (S&P)



IISER PUNE

## PRE-BID CONFERENCE FOR PROCUREMENT OF CRYOGENIC PROBE STATION WITH ACCESSORIES

### TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-1020-14

DATE : 03.12.14

| S.No | Query/Clarification Sought   | Clarification / Amendment   |
|------|--|---|
| 1    | <p><b>Page 23 : Point number 7</b></p> <p>Whether micromanipulated stages are automated or manual.</p>   | <p><b>Page 23 : Point number 7</b></p> <p>We have not mentioned automated in our specifications. If that option is available, bidders can quote it separately.</p> <p>There will be no change in specifications</p>   |
| 2    | <p><b>Page No -23 : Point number 12</b></p> <p>Magnetic Field produced by the Permanent Magnets and its direction ( horizontal or perpendicular or both)</p> | <p><b>Page No -23 : Point number 12 - has been modified as</b></p> <p>Option of variations in magnetic field using permanent magnets (upto 1-2 kOe) in vertical and horizontal direction is required. Permanent Magnets should be removable for zero field measurements.</p> <p>If system can be configured only for one field direction ( either vertical or horizontal ) then vertical direction will be preferred. Both the options can be quoted as accessories separately.</p> |



IISER PUNE

**PRE-BID CONFERENCE FOR PROCUREMENT OF CRYOGENIC PROBE STATION WITH ACCESSORIES**

## COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-1020-14

DATE : 3.12.14

| S.No | Query/Clarification Sought | Clarification / Amendment |
|------|----------------------------|---------------------------|
|      |                            |                           |
|      | -----NIL-----              | -----NIL-----             |
|      |                            |                           |
|      |                            |                           |
|      |                            |                           |